


LOT PX 446

CONTROL

 (Exp.) **2024-05-05**
(YYYY-MM-DD)

Rev 1

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
				DX NEXUS DF NEXUS					DX NEXUS DF NEXUS						DX NEXUS DF NEXUS				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l			3.0			± 0.4			8.8			± 1.0			19.3			± 2.2
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l			2.35			± 0.16			4.58			± 0.20			5.30			± 0.25
HB HGB	g/dl			6.1			± 0.4			13.3			± 0.5			16.5			± 0.6
	g/l			61			± 4			133			± 5			165			± 6
	mmol/l			3.79			± 0.25			8.26			± 0.31			10.25			± 0.37
HT HCT	%			18.6			± 1.5			38.9			± 2.0			48.2			± 2.5
	l/l			0.186			± 0.015			0.389			± 0.020			0.482			± 0.025
VGM MCV	µm ³ ; fl			79			± 5			85			± 5			91			± 5
TGMH MCH	pg			26.0			± 2.0			29.0			± 2.0			31.1			± 2.5
	fmol			1.61			± 0.12			1.80			± 0.12			1.93			± 0.16
CCMH MCHC	g/dl			32.9			± 3.0			34.2			± 3.0			34.2			± 3.0
	g/l			329			± 30			342			± 30			342			± 30
	mmol/l			20.40			± 1.86			21.22			± 1.86			21.25			± 1.86
IDR RDW	%			18.5			± 4.0			15.5			± 4.0			14.0			± 4.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l			68			± 20			236			± 30			453			± 50
VPM MPV	µm ³ ; fl			9.6			± 2.0			9.1			± 2.0			9.9			± 2.0
NEUT	#			1.59			± 0.35			5.06			± 0.90			13.30			± 1.90
	%			53.1			± 10.0			57.5			± 10.0			68.9			± 10.0
LYMPHO	#			0.89			± 0.33			2.71			± 0.70			3.61			± 1.50
	%			29.8			± 12.0			30.8			± 8.0			18.7			± 8.0
MONO	#			0.16			± 0.16			0.32			± 0.32			0.56			± 0.56
	%			5.3			± 5.3			3.6			± 3.6			2.9			± 2.9
EOS	#			0.21			± 0.21			0.30			± 0.30			0.95			± 0.95
	%			7.1			± 7.1			3.4			± 3.4			4.9			± 4.9
BASO	#			0.14			± 0.14			0.41			± 0.41			0.89			± 0.89
	%			4.7			± 4.7			4.7			± 4.7			4.6			± 4.6

Ref: TEMP-0821 Rev.50 BACK / VERSO