

# ABX Minotrol Retic



**LOT** RX 446  
Rev 1

**CONTROL**

(Exp.) 2024-05-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
				1				2				3	
				PENTRA			PENTRA			PENTRA			
		DX Nexus			DX Nexus			DX Nexus					
Méthode Manuelle / Manual RETIC	%			<b>1.6</b>	± 0.8			<b>5.4</b>	± 2.2			<b>10.4</b>	± 3.2
RETIC	%			<b>1.41</b>	± 0.71			<b>4.17</b>	± 1.04			<b>8.84</b>	± 1.77
RETIC	10 <sup>6</sup> /µl			<b>0.049</b>	± 0.025			<b>0.146</b>	± 0.037			<b>0.302</b>	± 0.060
MFI MRI *	%			<b>23.9</b>	± 8.0			<b>29.9</b>	± 10.0			<b>32.1</b>	± 12.0
GR RBC	10 <sup>6</sup> /µl			<b>3.50</b>	± 0.30			<b>3.49</b>	± 0.30			<b>3.42</b>	± 0.30
PIC RANGE	-			<b>27.0</b>	± 5.0			<b>27.0</b>	± 5.0			<b>27.0</b>	± 5.0

\* For research use only in Canada and USA

Ref: TEMP-0382 Rev.45 FRONT / RECTO 1300090305

# ABX Minotrol Retic



**LOT** RX 446  
Rev 1

**CONTROL**

(Exp.) 2024-05-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA				PENTRA				PENTRA			
XLR			XLR			XLR							
Méthode Manuelle / Manual RETIC	%	<b>1.6</b>			± 0.8	<b>5.4</b>			± 2.2	<b>10.4</b>			± 3.2
RETIC	%	<b>1.63</b>			± 0.82	<b>4.61</b>			± 1.15	<b>9.52</b>			± 1.90
RETIC	10 <sup>6</sup> /µl	<b>0.056</b>			± 0.028	<b>0.158</b>			± 0.040	<b>0.318</b>			± 0.064
MFI MRI *	%	<b>22.5</b>			± 8.0	<b>27.6</b>			± 10.0	<b>28.9</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.43</b>			± 0.30	<b>3.42</b>			± 0.30	<b>3.34</b>			± 0.30
PIC RANGE	-	<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.45 BACK / VERSO 1300090305

\* For research use only in Canada and USA