

# ABX Minocal



**LOT**

**CX 492**  
Rev 1

**CAL**

(Exp.) **2024-04-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP	LC-767G	MS CRP	XLR	DX NEXUS DF NEXUS	MS60		60 60 C+ ES60	XL80	
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>9.2</b>	<b>9.1</b>	<b>9.3</b>	<b>9.2</b>	<b>9.4</b>	<b>9.1</b>		<b>9.2</b>	<b>9.2</b>	± 0.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>4.47</b>	<b>4.46</b>	<b>4.48</b>	<b>4.50</b>	<b>4.51</b>	<b>4.48</b>		<b>4.50</b>	<b>4.50</b>	± 0.06
HB HGB	g/dl	<b>13.2</b>	<b>13.2</b>	<b>13.4</b>	<b>13.4</b>	<b>13.3</b>	<b>13.3</b>		<b>13.4</b>	<b>13.4</b>	± 0.2
	g/l	<b>132</b>	<b>132</b>	<b>134</b>	<b>134</b>	<b>133</b>	<b>133</b>		<b>134</b>	<b>134</b>	± 2
	mmol/l	<b>8.20</b>	<b>8.20</b>	<b>8.32</b>	<b>8.32</b>	<b>8.26</b>	<b>8.26</b>		<b>8.32</b>	<b>8.32</b>	± 0.12
HT HCT	%	<b>37.5</b>	<b>37.2</b>	<b>36.5</b>	<b>36.7</b>	<b>36.5</b>	<b>36.3</b>		<b>36.7</b>	<b>36.7</b>	± 1.0
	l/l	<b>0.375</b>	<b>0.372</b>	<b>0.365</b>	<b>0.367</b>	<b>0.365</b>	<b>0.363</b>		<b>0.367</b>	<b>0.367</b>	± 0.010
PLA PLT	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>233</b>	<b>239</b>	<b>222</b>	<b>226</b>	<b>235</b>	<b>227</b>		<b>229</b>	<b>226</b>	± 10
VMP MPV	µm <sup>3</sup> ; fl	<b>8.9</b>	<b>8.9</b>	N/A	N/A	N/A	N/A		N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.51 BACK / VERSO 1300090308