

ABX Minotrol Retic



LOT RX 446
Rev 1

CONTROL

(Exp.) 2024-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
				1				2				3	
				PENTRA				PENTRA				PENTRA	
		DX Nexus			DX Nexus			DX Nexus					
Méthode Manuelle / Manual RETIC	%			1.6	± 0.8			5.4	± 2.2			10.4	± 3.2
RETIC	%			1.41	± 0.71			4.17	± 1.04			8.84	± 1.77
RETIC	10 ⁶ /µl			0.049	± 0.025			0.146	± 0.037			0.302	± 0.060
MFI MRI *	%			23.9	± 8.0			29.9	± 10.0			32.1	± 12.0
GR RBC	10 ⁶ /µl			3.50	± 0.30			3.49	± 0.30			3.42	± 0.30
PIC RANGE	-			27.0	± 5.0			27.0	± 5.0			27.0	± 5.0

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.45 FRONT / RECTO 1300090305

ABX Minotrol Retic



LOT RX 446
Rev 1

CONTROL

(Exp.) 2024-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA				PENTRA				PENTRA			
XLR			XLR			XLR							
Méthode Manuelle / Manual RETIC	%	1.6			± 0.8	5.4			± 2.2	10.4			± 3.2
RETIC	%	1.63			± 0.82	4.61			± 1.15	9.52			± 1.90
RETIC	10 ⁶ /µl	0.056			± 0.028	0.158			± 0.040	0.318			± 0.064
MFI MRI *	%	22.5			± 8.0	27.6			± 10.0	28.9			± 12.0
GR RBC	10 ⁶ /µl	3.43			± 0.30	3.42			± 0.30	3.34			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.45 BACK / VERSO 1300090305

* For research use only in Canada and USA