

# ABX Minotrol CRP



**LOT**

**MC 431**

**CONTROL**



**(Exp.) 2024-07-05**  
(YYYY - MM - DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio											
				CONTROL			1	CONTROL			2	CONTROL			3
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE
					60 8P 60 18P				60 8P 60 18P				60 8P 60 18P		
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l		<b>2.1</b>		± 0.4		<b>7.8</b>		± 1.0		<b>20.8</b>		± 2.2	
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l		<b>2.24</b>		± 0.15		<b>4.52</b>		± 0.18		<b>5.57</b>		± 0.25	
HB	HGB	g/dl		<b>6.0</b>		± 0.4		<b>13.3</b>		± 0.5		<b>17.9</b>		± 0.7	
		g/l		<b>60</b>		± 4		<b>133</b>		± 5		<b>179</b>		± 7	
		mmol/l		<b>3.73</b>		± 0.25		<b>8.26</b>		± 0.31		<b>11.12</b>		± 0.44	
HT	HCT	%		<b>16.8</b>		± 2.0		<b>37.1</b>		± 2.5		<b>50.1</b>		± 3.0	
		l/l		<b>0.168</b>		± 0.020		<b>0.371</b>		± 0.025		<b>0.501</b>		± 0.030	
VGM	MCV	µm <sup>3</sup> ; fl		<b>75</b>		± 4		<b>82</b>		± 4		<b>90</b>		± 4	
TGMH	MCH	pg		<b>26.8</b>		± 2.4		<b>29.4</b>		± 2.4		<b>32.1</b>		± 2.8	
		fmol		<b>1.66</b>		± 0.15		<b>1.83</b>		± 0.15		<b>2.00</b>		± 0.18	
CCMH	MCHC	g/dl		<b>35.7</b>		± 3.0		<b>35.9</b>		± 3.0		<b>35.7</b>		± 3.0	
		g/l		<b>357</b>		± 30		<b>359</b>		± 30		<b>357</b>		± 30	
		mmol/l		<b>22.18</b>		± 1.86		<b>22.28</b>		± 1.86		<b>22.17</b>		± 1.86	
IDR	RDW	%		<b>14.5</b>		± 3.0		<b>14.5</b>		± 3.0		<b>14.0</b>		± 3.0	
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l		<b>80</b>		± 20		<b>274</b>		± 40		<b>536</b>		± 60	
VPM	MPV	µm <sup>3</sup> ; fl		<b>8.8</b>		± 2.0		<b>8.0</b>		± 2.0		<b>8.1</b>		± 2.0	
LYMPHO		#		<b>1.4</b>		± 0.2		<b>2.7</b>		± 0.5		<b>3.5</b>		± 1.1	
		%		<b>66.5</b>		± 8.0		<b>34.0</b>		± 6.0		<b>17.0</b>		± 5.0	
MONO		#		<b>0.2</b>		± 0.2		<b>1.2</b>		± 1.1		<b>1.7</b>		± 1.3	
		%		<b>8.0</b>		± 7.0		<b>15.3</b>		± 13.8		<b>8.0</b>		± 6.0	
GRANULO		#		<b>0.5</b>		± 0.2		<b>4.0</b>		± 1.0		<b>15.6</b>		± 2.1	
		%		<b>25.5</b>		± 7.0		<b>50.7</b>		± 10.0		<b>75.0</b>		± 10.0	
CRP		mg/dl		N/A		± 0.30		N/A		± 0.70		N/A		± 1.00	
		mg/l		N/A		± 3.0		N/A		± 7.0		N/A		± 10.0	

Ref: TEMP-1070 Rev.42 1300027267-A