

ABX Minotrol CRP



LOT

MC 431

CONTROL



(Exp.) 2024-07-05
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio														TOLERANCES TOLERANCE		
				CONTROL				1	CONTROL				2	CONTROL						3
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA				TOLERANCES TOLERANCE
				EMI CRP	LC-767G	MS CRP	EMI CRP		LC-767G	MS CRP	EMI CRP	LC-767G		MS CRP						
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.2	2.1	2.1		± 0.4	8.0	7.9	8.1		± 1.0	21.3	21.4	22.1		± 2.2			
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.30	2.30	2.20		± 0.15	4.50	4.48	4.60		± 0.18	5.50	5.45	5.60		± 0.25			
HB	HGB	g/dl	6.3	6.2	6.0		± 0.4	13.2	13.2	13.4		± 0.5	17.6	17.5	18.1		± 0.7			
		g/l	63	62	60		± 4	132	132	134		± 5	176	175	181		± 7			
		mmol/l	3.91	3.85	3.73		± 0.25	8.20	8.20	8.32		± 0.31	10.93	10.87	11.24		± 0.44			
HT	HCT	%	16.8	16.8	16.4		± 2.0	36.0	35.8	36.8		± 2.5	48.4	48.0	49.0		± 3.0			
		l/l	0.168	0.168	0.164		± 0.020	0.360	0.358	0.368		± 0.025	0.484	0.480	0.490		± 0.030			
VGM	MCV	µm ³ ; fl	73	73	74.5		± 4	80	80	80.0		± 4	88	88	87.5		± 4			
TGMH	MCH	pg	27.4	27.0	27.3		± 2.4	29.3	29.5	29.1		± 2.4	32.0	32.1	32.3		± 2.8			
		fmol	1.70	1.67	1.69		± 0.15	1.82	1.83	1.81		± 0.15	1.99	1.99	2.01		± 0.18			
CCMH	MCHC	g/dl	37.5	36.9	36.6		± 3.0	36.7	36.8	36.4		± 3.0	36.4	36.5	36.9		± 3.0			
		g/l	375	369	366		± 30	367	368	364		± 30	364	365	369		± 30			
		mmol/l	23.30	22.93	22.73		± 1.86	22.77	22.87	22.61		± 1.86	22.58	22.66	22.94		± 1.86			
IDR	RDW	%	13.5	13.0	15.5		± 3.0	14.0	13.5	13.0		± 3.0	13.5	13.0	12.0		± 3.0			
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	77	73	73		± 20	256	257	271		± 40	500	509	527		± 60			
VPM	MPV	µm ³ ; fl	7.5	8.4	8.9		± 2.0	7.5	7.8	8.3		± 2.0	7.7	7.9	8.4		± 2.0			
LYMPHO		#	1.4	1.3	N/A		± 0.2	2.9	2.7	N/A		± 0.5	3.6	3.4	N/A		± 1.1			
		%	64.5	64.0	N/A		± 8.0	36.0	34.5	N/A		± 6.0	17.0	16.0	N/A		± 5.0			
MONO		#	0.2	0.2	N/A		± 0.2	1.4	1.2	N/A		± 1.1	2.8	2.4	N/A		± 1.8			
		%	8.0	8.0	N/A		± 7.0	17.3	15.0	N/A		± 13.5	13.0	11.0	N/A		± 8.0			
GRANULO		#	0.6	0.6	N/A		± 0.2	3.7	4.0	N/A		± 1.0	14.9	15.6	N/A		± 2.1			
		%	27.5	28.0	N/A		± 7.0	46.7	50.5	N/A		± 10.0	70.0	73.0	N/A		± 10.0			
CRP		mg/dl	0.54	0.51	0.55		± 0.40	2.52	2.60	2.63		± 0.80	4.75	4.88	5.07		± 1.50			
		mg/l	5.4	5.1	5.5		± 4.0	25.2	26.0	26.3		± 8.0	47.5	48.8	50.7		± 15.0			

Ref: TEMP-1070 Rev.42 BACKVERSO 1300092170