

ABX Minotrol CRP



LOT

MC 431

CONTROL

(Exp.) **2024-07-05**
(YYYY - MM - DD)

Rev 2

		ABX Lyse																				
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					1	CONTROL					2	CONTROL					3			
		MICROS					TOLERANCES TOLERANCE	MICROS					TOLERANCES TOLERANCE	MICROS					TOLERANCES TOLERANCE			
				60 8P 60 18P						60 8P 60 18P						60 8P 60 18P						
GB WBC	10 ³ /mm ³ ;10 ⁹ /l			2.1			± 0.4			7.9					± 1.0			21.1				± 2.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l			2.25			± 0.15			4.54					± 0.18			5.59				± 0.25
HB HGB	g/dl			6.2			± 0.4			13.4					± 0.5			17.7				± 0.7
	g/l			62			± 4			134					± 5			177				± 7
	mmol/l			3.85			± 0.25			8.32					± 0.31			10.99				± 0.44
HT HCT	%			16.7			± 2.0			36.8					± 2.5			49.8				± 3.0
	l/l			0.167			± 0.020			0.368					± 0.025			0.498				± 0.030
VGM MCV	µm ³ ; fl			74			± 4			81					± 4			89				± 4
TGMH MCH	pg			27.6			± 2.4			29.5					± 2.4			31.7				± 2.8
	fmol			1.71			± 0.15			1.83					± 0.15			1.97				± 0.18
CCMH MCHC	g/dl			37.2			± 3.0			36.4					± 3.0			35.6				± 3.0
	g/l			372			± 30			364					± 30			356				± 30
	mmol/l			23.12			± 1.86			22.63					± 1.86			22.09				± 1.86
IDR RDW	%			14.0			± 3.0			14.0					± 3.0			13.0				± 3.0
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l			80			± 20			268					± 40			532				± 60
VPM MPV	µm ³ ; fl			7.5			± 2.0			6.8					± 2.0			7.1				± 2.0
LYMPHO	#			1.3			± 0.2			2.7					± 0.5			3.3				± 1.1
	%			64.0			± 8.0			34.0					± 6.0			15.5				± 5.0
MONO	#			0.2			± 0.2			1.2					± 1.1			1.7				± 1.3
	%			10.0			± 7.0			15.3					± 13.8			8.0				± 6.0
GRANULO	#			0.5			± 0.2			4.0					± 1.0			16.1				± 2.1
	%			26.0			± 7.0			50.7					± 10.0			76.5				± 10.0
CRP	mg/dl			N/A			± 0.30			N/A					± 0.70			N/A				± 1.00
	mg/l			N/A			± 3.0			N/A					± 7.0			N/A				± 10.0

Ref: TEMP-1070 Rev.42 FRONT/RECTO 1300092170