

ABX Minocal

LOT

CX 494

CAL



(Exp.) 2024-06-05
(YYYY-MM-DD)

Rev 1

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|---|--------------|--------|--------|--------|-----------------|--------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | PENTRA | PENTRA | PENTRA | | PENTRA |
| | | ES60 | | | | 60 8P 60 18P | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 8.9 | | | | 9.1 | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.43 | | | | 4.42 | | | | | ± 0.06 |
| HB HGB | g/dl | 13.3 | | | | 13.3 | | | | | ± 0.2 |
| | g/l | 133 | | | | 133 | | | | | ± 2 |
| | mmol/l | 8.26 | | | | 8.26 | | | | | ± 0.12 |
| HT HCT | % | 37.4 | | | | 37.6 | | | | | ± 1.0 |
| | l/l | 0.374 | | | | 0.376 | | | | | ± 0.010 |
| PLA PLT | 10 ³ /mm ³ ;10 ⁹ /l | 226 | | | | 237 | | | | | ± 10 |
| VMP MPV | μm ³ ; fl | 8.8 | | | | 9.2 | | | | | ± 0.5 |

ABX Minocal

LOT

CX 494

Rev 1

CAL



(Exp.) 2024-06-05
(YYYY-MM-DD)

| PARAMETRES PARAMETERS | UNITES UNITS | LyseVet | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|---------|-----------------------------|--|--|--|--|--|--|-------------------------|
| | | MICROS | MICROS | | | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | | Scil Vet ABC Plus ESV 60 | | | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | | | | | | | | | ± 0.06 |
| HB HGB | g/dl | | | | | | | | | ± 0.2 |
| | g/l | | | | | | | | | ± 2 |
| | mmol/l | | | | | | | | | ± 0.12 |
| HT HCT | % | | | | | | | | | ± 1.0 |
| | l/l | | | | | | | | | ± 0.010 |
| PLA PLT | 10 ³ /mm ³ ;10 ⁹ /l | | | | | | | | | ± 10 |
| VMP MPV | µm ³ ; fl | | | | | | | | | ± 0.5 |

Ref: TEMP-0387 Rev.51 FRONT / RECTO 1300113555