

LOT MX 447 L
Rev 1

CONTROL L

 (Exp.)

2024-08-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60				60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0				2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.23				2.26					± 0.17
HB HGB	g/dl	5.8				5.7					± 0.5
	g/l	58				57					± 5
	mmol/l	3.60				3.52					± 0.31
HT HCT	%	16.1				16.5					± 2.0
	l/l	0.161				0.165					± 0.020
VGM MCV	µm ³ ; fl	72				73					± 4
TGMH MCH	pg	26.0				25.1					± 2.0
	fmol	1.62				1.56					± 0.12
CCMH MCHC	g/dl	36.1				34.4					± 3.0
	g/l	361				344					± 30
	mmol/l	22.43				21.35					± 1.86
IDR RDW	%	15.0				15.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	72				79					± 22
VPM MPV	µm ³ ; fl	8.4				7.4					± 2.0
LYMPHO	#	1.24				1.28					± 0.40
	%	62.0				64.0					± 8.0
MONO	#	0.20				0.15					± 0.20
	%	8.5				7.5					± 6.0
GRANULO	#	0.59				0.57					± 0.40
	%	29.5				28.5					± 7.0

LOT MX 447 N
Rev 1

CONTROL N

(Exp.)

2024-08-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60				60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.5				7.3					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.51				4.53					± 0.26
HB HGB	g/dl	13.4				13.2					± 0.7
	g/l	134				132					± 7
	mmol/l	8.32				8.22					± 0.43
HT HCT	%	37.0				37.1					± 2.7
	l/l	0.370				0.371					± 0.027
VGM MCV	µm ³ ; fl	82				82					± 4
TGMH MCH	pg	29.7				29.2					± 2.0
	fmol	1.85				1.81					± 0.12
CCMH MCHC	g/dl	36.2				35.6					± 3.0
	g/l	362				356					± 30
	mmol/l	22.50				22.12					± 1.86
IDR RDW	%	15.0				15.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	244				259					± 40
VPM MPV	µm ³ ; fl	7.8				6.7					± 2.0
LYMPHO	#	2.59				2.63					± 0.40
	%	34.5				36.0					± 5.0
MONO	#	0.56				0.47					± 0.45
	%	7.5				6.5					± 6.0
GRANULO	#	4.35				4.20					± 0.60
	%	58.0				57.5					± 6.0

Ref: TEMP-0830 Rev.43 FRONT / RECTO 1300104143

LOT MX 447 H
Rev 1

CONTROL H

 (Exp.)

2024-08-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60				60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.3				19.7					± 1.7
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.76				5.70					± 0.32
HB HGB	g/dl	18.6				18.1					± 0.9
	g/l	186				181					± 9
	mmol/l	11.55				11.25					± 0.56
HT HCT	%	51.3				51.3					± 3.4
	l/l	0.513				0.513					± 0.034
VGM MCV	µm ³ ; fl	89				90					± 4
TGMH MCH	pg	32.3				31.8					± 2.0
	fmol	2.01				1.97					± 0.12
CCMH MCHC	g/dl	36.3				35.3					± 3.0
	g/l	363				353					± 30
	mmol/l	22.53				21.93					± 1.86
IDR RDW	%	14.5				14.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	478				494					± 65
VPM MPV	µm ³ ; fl	7.6				6.6					± 2.0
LYMPHO	#	3.25				3.05					± 1.00
	%	16.0				15.5					± 5.0
MONO	#	1.02				0.99					± 0.80
	%	5.0				5.0					± 4.0
GRANULO	#	16.04				15.66					± 1.40
	%	79.0				79.5					± 7.0