


LOT MX 447 L
Rev 1

CONTROL L

 (Exp.)

2024-08-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE							
		MICROS															
		EMI CRP															
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l									2.0							± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l									2.29							± 0.15
HB HGB	g/dl									6.1							± 0.4
	g/l									61							± 4
	mmol/l									3.79							± 0.25
HT HCT	%									16.3							± 2.0
	l/l									0.163							± 0.020
VGM MCV	µm ³ ; fl									71							± 4
TGMH MCH	pg									26.6							± 2.0
	fmol									1.65							± 0.12
CCMH MCHC	g/dl									37.5							± 3.0
	g/l									375							± 30
	mmol/l									23.30							± 1.86
IDR RDW	%									14.5							± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l									74							± 20
VPM MPV	µm ³ ; fl									7.4							± 2.0
LYMPHO	#									1.23							± 0.40
	%									61.5							± 8.0
MONO	#									0.20							± 0.20
	%									8.0							± 6.0
GRANULO	#									0.61							± 0.40
	%									30.5							± 7.0

LOT		MX 447 N	CONTROL		N					 (Exp.)	2024-08-05 (YYYY-MM-DD)	
		Rev 1	ABX Lysebio									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE	
				EMI CRP								
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l			7.7							± 0.8	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l			4.50							± 0.18	
HB HGB	g/dl			13.4							± 0.5	
	g/l			134							± 5	
	mmol/l			8.32							± 0.31	
HT HCT	%			36.5							± 2.5	
	l/l			0.365							± 0.025	
VGM MCV	µm ³ ; fl			81							± 4	
TGMH MCH	pg			29.8							± 2.0	
	fmol			1.85							± 0.12	
CCMH MCHC	g/dl			36.8							± 3.0	
	g/l			368							± 30	
	mmol/l			22.83							± 1.86	
IDR RDW	%			14.0							± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l			257							± 40	
VPM MPV	µm ³ ; fl			7.5							± 2.0	
LYMPHO	#			2.70							± 0.40	
	%			35.0							± 5.0	
MONO	#			0.58							± 0.45	
	%			7.5							± 6.0	
GRANULO	#			4.43							± 0.60	
	%			57.5							± 6.0	

Ret: TEMP-0830 Rev.43 BACK / VERSO 1300104144

LOT

MX 447 H
Rev 1

CONTROL

H

 (Exp.)

2024-08-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE
				MICROS EMI CRP						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l			20.5						± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l			5.67						± 0.20
HB HGB	g/dl			18.0						± 0.6
	g/l			180						± 6
	mmol/l			11.18						± 0.37
HT HCT	%			49.9						± 3.0
	l/l			0.499						± 0.030
VGM MCV	µm ³ ; fl			88						± 4
TGMH MCH	pg			31.7						± 2.0
	fmol			1.97						± 0.12
CCMH MCHC	g/dl			36.1						± 3.0
	g/l			361						± 30
	mmol/l			22.40						± 1.86
IDR RDW	%			13.5						± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l			481						± 45
VPM MPV	µm ³ ; fl			7.4						± 2.0
LYMPHO	#			3.38						± 1.00
	%			16.5						± 5.0
MONO	#			1.33						± 0.80
	%			6.5						± 4.0
GRANULO	#			15.79						± 1.40
	%			77.0						± 7.0