


LOT PX 447

CONTROL

 (Exp.) **2024-07-05**
(YYYY-MM-DD)

Rev 1

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
				DX NEXUS DF NEXUS					DX NEXUS DF NEXUS						DX NEXUS DF NEXUS				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l			3.1			± 0.4			9.1			± 1.0			19.3			± 2.2
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l			2.33			± 0.16			4.59			± 0.20			5.23			± 0.25
HB HGB	g/dl			6.2			± 0.4			13.4			± 0.5			16.5			± 0.6
	g/l			62			± 4			134			± 5			165			± 6
	mmol/l			3.85			± 0.25			8.32			± 0.31			10.25			± 0.37
HT HCT	%			18.6			± 1.5			39.5			± 2.0			48.1			± 2.5
	l/l			0.186			± 0.015			0.395			± 0.020			0.481			± 0.025
VGM MCV	µm ³ ; fl			80			± 5			86			± 5			92			± 5
TGMH MCH	pg			26.6			± 2.0			29.2			± 2.0			31.5			± 2.5
	fmol			1.65			± 0.12			1.81			± 0.12			1.96			± 0.16
CCMH MCHC	g/dl			33.3			± 3.0			33.9			± 3.0			34.3			± 3.0
	g/l			333			± 30			339			± 30			343			± 30
	mmol/l			20.66			± 1.86			21.08			± 1.86			21.30			± 1.86
IDR RDW	%			17.0			± 4.0			15.5			± 4.0			13.5			± 4.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l			66			± 20			254			± 30			474			± 50
VPM MPV	µm ³ ; fl			8.7			± 2.0			8.2			± 2.0			8.8			± 2.0
NEUT	#			1.61			± 0.35			4.92			± 0.90			13.88			± 1.90
	%			51.9			± 10.0			54.1			± 10.0			71.9			± 10.0
LYMPHO	#			1.01			± 0.33			3.16			± 0.70			3.15			± 1.50
	%			32.5			± 12.0			34.7			± 8.0			16.3			± 8.0
MONO	#			0.11			± 0.11			0.24			± 0.24			0.42			± 0.42
	%			3.6			± 3.6			2.6			± 2.6			2.2			± 2.2
EOS	#			0.23			± 0.23			0.35			± 0.35			0.87			± 0.87
	%			7.4			± 7.4			3.8			± 3.8			4.5			± 4.5
BASO	#			0.14			± 0.14			0.44			± 0.44			0.98			± 0.98
	%			4.6			± 4.6			4.8			± 4.8			5.1			± 5.1

Ref: TEMP-0821 Rev.50 BACK / VERSO