

ABX Minotrol Retic



LOT RX 447
Rev 1

CONTROL

(Exp.) 2024-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
				1				2				3	
				PENTRA				PENTRA				PENTRA	
		DX Nexus			DX Nexus			DX Nexus					
Méthode Manuelle / Manual RETIC	%			1.4	± 0.7			4.8	± 2.0			10.7	± 3.3
RETIC	%			1.56	± 0.78			4.53	± 1.13			9.31	± 1.86
RETIC	10 ⁶ /µl			0.053	± 0.027			0.156	± 0.039			0.321	± 0.064
MFI MRI *	%			24.0	± 8.0			30.3	± 10.0			31.9	± 12.0
GR RBC	10 ⁶ /µl			3.37	± 0.30			3.44	± 0.30			3.45	± 0.30
PIC RANGE	-			27.0	± 5.0			27.0	± 5.0			27.0	± 5.0

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.45 FRONT / RECTO 1300090305

ABX Minotrol Retic



LOT RX 447
Rev 1

CONTROL

(Exp.) 2024-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITS UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA				PENTRA				PENTRA			
XLR			XLR			XLR							
Méthode Manuelle / Manual RETIC	%	1.4			± 0.7	4.8			± 2.0	10.7			± 3.3
RETIC	%	1.81			± 0.91	4.91			± 1.23	10.04			± 2.01
RETIC	10 ⁶ /µl	0.060			± 0.030	0.166			± 0.042	0.342			± 0.068
MFI MRI *	%	23.5			± 8.0	29.1			± 10.0	30.3			± 12.0
GR RBC	10 ⁶ /µl	3.33			± 0.30	3.39			± 0.30	3.41			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.45 BACK / VERSO 1300090305

* For research use only in Canada and USA