

LOT PX 447
Rev 1


CONTROL

 (Exp.) 2024-07-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		XL80	XLR					XL80	XLR					XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.2	3.2			± 0.4	8.9	8.9			± 1.0	18.7	18.7			± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.29	2.29			± 0.16	4.53	4.53			± 0.20	5.12	5.12			± 0.25			
HB HGB	g/dl	6.1	6.1			± 0.4	13.4	13.4			± 0.5	16.6	16.6			± 0.6			
	g/l	61	61			± 4	134	134			± 5	166	166			± 6			
	mmol/l	3.79	3.79			± 0.25	8.32	8.32			± 0.31	10.31	10.31			± 0.37			
HT HCT	%	18.1	18.1			± 1.5	38.5	38.5			± 2.0	47.1	47.1			± 2.5			
	l/l	0.181	0.181			± 0.015	0.385	0.385			± 0.020	0.471	0.471			± 0.025			
VGM MCV	µm ³ fl	79	79			± 5	85	85			± 5	92	92			± 5			
TGMH MCH	pg	26.6	26.6			± 2.0	29.6	29.6			± 2.0	32.4	32.4			± 2.5			
	fmol	1.65	1.65			± 0.12	1.84	1.84			± 0.12	2.01	2.01			± 0.16			
CCMH MCHC	g/dl	33.7	33.7			± 3.0	34.8	34.8			± 3.0	35.2	35.2			± 3.0			
	g/l	337	337			± 30	348	348			± 30	352	352			± 30			
	mmol/l	20.94	20.94			± 1.86	21.61	21.61			± 1.86	21.88	21.88			± 1.86			
IDR RDW	%	14.5	14.5			± 4.0	13.5	13.5			± 4.0	12.0	12.0			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	61	61			± 20	253	253			± 30	470	470			± 50			
VPM MPV	µm ³ fl	8.4	8.4			± 2.0	8.4	8.4			± 2.0	9.1	9.1			± 2.0			
NEUT	#	1.50	1.50			± 0.35	4.58	4.58			± 0.90	13.26	13.26			± 1.90			
	%	47.0	47.0			± 10.0	51.5	51.5			± 10.0	70.9	70.9			± 10.0			
LYMPHO	#	1.31	1.31			± 0.33	3.57	3.57			± 0.70	3.57	3.57			± 1.50			
	%	40.9	40.9			± 12.0	40.1	40.1			± 8.0	19.1	19.1			± 8.0			
MONO	#	0.11	0.11			± 0.11	0.21	0.21			± 0.21	0.39	0.39			± 0.39			
	%	3.4	3.4			± 3.4	2.4	2.4			± 2.4	2.1	2.1			± 2.1			
EOS	#	0.20	0.20			± 0.20	0.28	0.28			± 0.28	0.75	0.75			± 0.75			
	%	6.1	6.1			± 6.1	3.2	3.2			± 3.2	4.0	4.0			± 4.0			
BASO	#	0.08	0.08			± 0.08	0.25	0.25			± 0.25	0.73	0.73			± 0.73			
	%	2.6	2.6			± 2.6	2.8	2.8			± 2.8	3.9	3.9			± 3.9			

LOT PX 447
Rev 1

CONTROL

 (Exp.) 2024-07-05
(YYYY - MM - DD)

		ABX Lysebio																		
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA						
		60 60C+ ES60	MS60	MSCR				60 60C+ ES60	MS60	MSCR				60 60C+ ES60	MS60	MSCR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.1	3.1	3.1			± 0.4	9.0	8.8	8.8			± 1.0	19.1	18.8	18.7			± 2.2	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.35	2.32	2.29			± 0.16	4.54	4.53	4.47			± 0.20	5.12	5.11	4.99			± 0.25	
	g/dl	6.1	6.1	6.1			± 0.4	13.4	13.4	13.3			± 0.5	16.5	16.5	16.4			± 0.6	
HB HGB	g/l	61	61	61			± 4	134	134	133			± 5	165	165	164			± 6	
	mmol/l	3.79	3.79	3.79			± 0.25	8.32	8.32	8.26			± 0.31	10.25	10.25	10.18			± 0.37	
HT HCT	%	18.1	17.9	17.6			± 1.5	38.6	38.1	37.8			± 2.0	47.6	47.0	45.9			± 2.5	
	l/l	0.181	0.179	0.176			± 0.015	0.386	0.381	0.378			± 0.020	0.476	0.470	0.459			± 0.025	
VGM MCV	µm ³ ; fl	77	77	77.0			± 5	85	84	84.5			± 5	93	92	92.0			± 5	
TGMH MCH	pg	26.0	26.3	26.6			± 2.0	29.5	29.6	29.8			± 2.0	32.2	32.3	32.9			± 2.5	
	fmol	1.61	1.63	1.65			± 0.12	1.83	1.84	1.85			± 0.12	2.00	2.01	2.04			± 0.16	
	g/dl	33.7	34.1	34.6			± 3.0	34.7	35.2	35.2			± 3.0	34.7	35.1	35.7			± 3.0	
CCMH MCHC	g/l	337	341	346			± 30	347	352	352			± 30	347	351	357			± 30	
	mmol/l	20.93	21.21	21.48			± 1.86	21.56	21.87	21.87			± 1.86	21.52	21.80	22.18			± 1.86	
IDR RDW	%	14.5	14.5	13.5			± 4.0	13.0	13.0	11.5			± 4.0	12.0	11.5	10.5			± 4.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	67	62	61			± 20	252	247	232			± 30	469	467	429			± 50	
VPM MPV	µm ³ ; fl	8.0	8.3	8.1			± 2.0	7.9	7.9	7.7			± 2.0	8.5	8.5	8.2			± 2.0	
NEUT	#	1.44	1.43	1.41			± 0.35	4.57	4.57	4.49			± 0.90	13.39	13.35	13.18			± 1.90	
	%	46.4	46.0	45.5			± 10.0	50.8	51.9	51.0			± 10.0	70.1	71.0	70.5			± 10.0	
LYMPHO	#	1.28	1.28	1.29			± 0.33	3.65	3.49	3.52			± 0.70	3.76	3.63	3.55			± 1.50	
	%	41.3	41.2	41.5			± 12.0	40.6	39.7	40.0			± 8.0	19.7	19.3	19.0			± 8.0	
MONO	#	0.11	0.11	0.12			± 0.11	0.23	0.19	0.22			± 0.19	0.44	0.34	0.37			± 0.34	
	%	3.7	3.4	4.0			± 3.4	2.6	2.2	2.5			± 2.2	2.3	1.8	2.0			± 1.8	
EOS	#	0.19	0.21	0.19			± 0.19	0.29	0.30	0.31			± 0.29	0.76	0.75	0.84			± 0.75	
	%	6.0	6.7	6.0			± 6.0	3.2	3.4	3.5			± 3.2	4.0	4.0	4.5			± 4.0	
BASO	#	0.08	0.08	0.09			± 0.08	0.25	0.25	0.26			± 0.25	0.74	0.73	0.75			± 0.73	
	%	2.6	2.7	3.0			± 2.6	2.8	2.8	3.0			± 2.8	3.9	3.9	4.0			± 3.9	

Ref: TEMP-0821 Rev.50 FRONT / RECTO