

# ABX Minotrol Retic



**LOT** RX 447  
Rev 1

**CONTROL**

(Exp.) 2024-07-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
				1				2				3	
				PENTRA				PENTRA				PENTRA	
		DX Nexus			DX Nexus			DX Nexus					
Méthode Manuelle / Manual RETIC	%			<b>1.4</b>	± 0.7			<b>4.8</b>	± 2.0			<b>10.7</b>	± 3.3
RETIC	%			<b>1.56</b>	± 0.78			<b>4.53</b>	± 1.13			<b>9.31</b>	± 1.86
RETIC	10 <sup>6</sup> /µl			<b>0.053</b>	± 0.027			<b>0.156</b>	± 0.039			<b>0.321</b>	± 0.064
MFI MRI *	%			<b>24.0</b>	± 8.0			<b>30.3</b>	± 10.0			<b>31.9</b>	± 12.0
GR RBC	10 <sup>6</sup> /µl			<b>3.37</b>	± 0.30			<b>3.44</b>	± 0.30			<b>3.45</b>	± 0.30
PIC RANGE	-			<b>27.0</b>	± 5.0			<b>27.0</b>	± 5.0			<b>27.0</b>	± 5.0

\* For research use only in Canada and USA

Ref: TEMP-0382 Rev.45 FRONT / RECTO 1300090305

# ABX Minotrol Retic



**LOT** RX 447  
Rev 1

**CONTROL**

(Exp.) 2024-07-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA		XLR		PENTRA		XLR		PENTRA		XLR	
Méthode Manuelle / Manual RETIC	%	<b>1.4</b>			± 0.7	<b>4.8</b>			± 2.0	<b>10.7</b>			± 3.3
RETIC	%	<b>1.81</b>			± 0.91	<b>4.91</b>			± 1.23	<b>10.04</b>			± 2.01
RETIC	10 <sup>6</sup> /µl	<b>0.060</b>			± 0.030	<b>0.166</b>			± 0.042	<b>0.342</b>			± 0.068
MFI MRI *	%	<b>23.5</b>			± 8.0	<b>29.1</b>			± 10.0	<b>30.3</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.33</b>			± 0.30	<b>3.39</b>			± 0.30	<b>3.41</b>			± 0.30
PIC RANGE	-	<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.45 BACK / VERSO 1300090305

\* For research use only in Canada and USA