


LOT PX 447
Rev 1

CONTROL

 (Exp.) 2024-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	Lysebio									
			CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
			YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
			H1500 HELO 1.x	H2500 HELO 1.x		H1500 HELO 1.x	H2500 HELO 1.x		H1500 HELO 1.x	H2500 HELO 1.x		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.55	2.55	± 0.40		7.50	7.50		± 1.00	17.70		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.40	2.40	± 0.16	4.61	4.61	± 0.20	5.19	5.19	± 0.25		
HB HGB	g/dl	6.2	6.2	± 0.4	13.5	13.5	± 0.5	16.5	16.5	± 0.6		
	g/l	62	62	± 4	135	135	± 5	165	165	± 6		
	mmol/l	3.85	3.85	± 0.25	8.38	8.38	± 0.31	10.25	10.25	± 0.37		
HT HCT	%	18.5	18.5	± 1.5	39.4	39.4	± 2.0	48.8	48.8	± 2.5		
	l/l	0.185	0.185	± 0.015	0.394	0.394	± 0.020	0.488	0.488	± 0.025		
VGM MCV	µm ³ ; fl	77.0	77.0	± 5.0	85.5	85.5	± 5.0	94.0	94.0	± 5.0		
TGMH MCH	pg	25.8	25.8	± 2.0	29.3	29.3	± 2.0	31.8	31.8	± 2.5		
	fmol	1.60	1.60	± 0.12	1.82	1.82	± 0.12	1.97	1.97	± 0.16		
CCMH MCHC	g/dl	33.5	33.5	± 3.0	34.3	34.3	± 3.0	33.8	33.8	± 3.0		
	g/l	335	335	± 30	343	343	± 30	338	338	± 30		
	mmol/l	20.83	20.83	± 1.86	21.27	21.27	± 1.86	21.00	21.00	± 1.86		
IDR-SD RDW-SD	fl	40.5	40.5	± 8.0	39.0	39.0	± 8.0	40.0	40.0	± 8.0		
IDR-CV RDW-CV	%	14.0	14.0	± 4.0	12.5	12.5	± 4.0	11.0	11.0	± 4.0		
PLA PLT	10 ⁹ /mm ³ ; 10 ⁹ /l	64	64	± 20	254	254	± 30	477	477	± 50		
PLT-Ox	10 ⁹ /mm ³ ; 10 ⁹ /l	N/A	65	± 30	N/A	250	± 40	N/A	484	± 50		
VMP MPV	µm ³ ; fl	8.8	8.8	± 2.0	8.6	8.6	± 2.0	9.4	9.4	± 2.0		
NEU	10 ⁹ /mm ³ ; 10 ⁹ /l	1.30	1.30	± 0.35	4.10	4.10	± 0.90	12.78	12.78	± 1.90		
	%	51.1	51.1	± 10.0	54.7	54.7	± 10.0	72.2	72.2	± 10.0		
LYM	10 ⁹ /mm ³ ; 10 ⁹ /l	0.76	0.76	± 0.33	2.39	2.39	± 0.70	2.64	2.64	± 1.50		
	%	29.9	29.9	± 12.0	31.9	31.9	± 8.0	14.9	14.9	± 8.0		
MON	10 ⁹ /mm ³ ; 10 ⁹ /l	0.20	0.20	± 0.20	0.44	0.44	± 0.40	0.76	0.76	± 0.68		
	%	7.7	7.7	± 7.70	5.8	5.8	± 5.2	4.3	4.3	± 3.9		
EOS	10 ⁹ /mm ³ ; 10 ⁹ /l	0.19	0.19	± 0.19	0.29	0.29	± 0.29	0.76	0.76	± 0.76		
	%	7.4	7.4	± 7.40	3.8	3.8	± 3.8	4.3	4.3	± 4.3		
BAS	10 ⁹ /mm ³ ; 10 ⁹ /l	0.10	0.10	± 0.10	0.29	0.29	± 0.19	0.76	0.76	± 0.38		
	%	3.9	3.9	± 2.5	3.8	3.8	± 3.0	4.3	4.3	± 2.5		
ERB NRBC	10 ⁹ /mm ³ ; 10 ⁹ /l	0.45	0.45	± 0.30	1.44	1.44	± 0.45	1.22	1.22	± 0.35		
	%	17.8	17.8	± 8.0	19.2	19.2	± 8.0	6.9	6.9	± 2.0		

BACK / VERSO
Ref: TEMP-0821 Rev.50