



## ***LAtrol™ Normal Control Plasma*** ***LAtrol®***

<b>HORIBA Reference No.:</b>	<b>1300079435</b>
<b>Biomedica Reference No.:</b>	<b>816N</b>
<b>Lot No.:</b>	<b>230110</b>
<b>Expiry date:</b>	<b>2025/01/10</b>

### **for Yumizen G800**

Parameter	Unit	Range
<b>DRVVT</b>	<b>sec</b>	<b>35.9 – 46.7</b>
<b>DRVVC</b>	<b>sec</b>	<b>33.8 – 43.4</b>
<b>Normalized Ratio</b>		<b>&lt;1.16</b>

### **for Yumizen G1500/G1550**

Parameter	Unit	Range
<b>DRVVT</b>	<b>sec</b>	<b>33.5 – 47.1</b>
<b>DRVVC</b>	<b>sec</b>	<b>33.4 – 46.8</b>
<b>Normalized Ratio</b>		<b>&lt;1.16</b>

#### **Important notice**

Each laboratory should establish its own DRVVT and DRVVC normal reference range (mean value ± 2 S.D.).

The Normalized Ratio is calculated as follows:

$$\text{Normalized Ratio} = \frac{\text{DRVVT (sample)} / \text{DRVVT (mean clot time test)}}{\text{DRVVC (sample)} / \text{DRVVC (mean clot time confirm)}}$$