



Yumizen G APTT Liq 4 for YUMIZEN G Line



Reference No.:		1300036381
Lot No.:		940202 (rev. 1)
Expiry date:		2025-08-31
MNAPTT		
Coagulometers		Sec
Optical	Yumizen G Line	29,3
Important notice		
<p>The Mean Normal Activated Partial Thromboplastin Time (MNAPTT) is for information only. The value depends on the population, race, gender, sampling tube, etc. According to the CLSI every laboratory should determine its local MNAPTT.</p>		