

# Yumizen G DDi 2 for YUMIZEN G1500/G1550



|                       |                            |                        |            |
|-----------------------|----------------------------|------------------------|------------|
| <b>Reference No.:</b> |                            | <b>1300036391</b>      |            |
| <b>Lot No.:</b>       |                            | <b>940110 (rev. 1)</b> |            |
| <b>Expiry date:</b>   |                            | <b>2025-01-31</b>      |            |
| <b>Coagulometers</b>  |                            | <b>MASTER CURVE</b>    |            |
| <b>Optical</b>        | <b>Yumizen G1500/G1550</b> | <b>µg(FEU)/mL</b>      | <b>ΔOD</b> |
|                       |                            | 4,00                   | 0.2785     |
|                       |                            | 3,00                   | 0.2140     |
|                       |                            | 2,00                   | 0.1290     |
|                       |                            | 1,00                   | 0.0535     |
|                       |                            | 0,50                   | 0.0248     |
|                       |                            | 0,25                   | 0.0122     |
| <b>Slope value</b>    |                            | <b>0,07</b>            |            |

# Yumizen G DDi 2 for YUMIZEN G800



|                       |                     |                        |            |
|-----------------------|---------------------|------------------------|------------|
| <b>Reference No.:</b> |                     | <b>1300036391</b>      |            |
| <b>Lot No.:</b>       |                     | <b>940110 (rev. 1)</b> |            |
| <b>Expiry date:</b>   |                     | <b>2025-01-31</b>      |            |
| <b>Coagulometers</b>  |                     | <b>MASTER CURVE</b>    |            |
| <b>Optical</b>        | <b>Yumizen G800</b> | <b>µg(FEU)/mL</b>      | <b>ΔOD</b> |
|                       |                     | 4,00                   | 0.2600     |
|                       |                     | 3,00                   | 0.2007     |
|                       |                     | 2,00                   | 0.1260     |
|                       |                     | 1,00                   | 0.0548     |
|                       |                     | 0,50                   | 0.0260     |
|                       |                     | 0,25                   | 0.0130     |
| <b>Slope value</b>    |                     | <b>0,06</b>            |            |

# Yumizen G DDi 2 for YUMIZEN G400 DDi/G200

|   |                              |                        |               |
|---|------------------------------|------------------------|---------------|
| <b>Reference No.:</b>   |                              | <b>1300036391</b>      |               |
| <b>Lot No.:</b>   |                              | <b>940110 (rev. 1)</b> |               |
| <b>Expiry date:</b>   |                              | <b>2025-01-31</b>      |               |
| <b>Coagulometers</b>  |                              | <b>MASTER CURVE</b>    |               |
|   |                              | <b>µg(FEU)/mL</b>      | <b>ΔOD</b>    |
| <b>Optical</b>  | <b>Yumizen G400 DDi/G200</b> | <b>4,00</b>            | <b>0.1030</b> |
|   |                              | <b>3,00</b>            | <b>0.0752</b> |
|   |                              | <b>2,00</b>            | <b>0.0475</b> |
|   |                              | <b>1,00</b>            | <b>0.0227</b> |
|   |                              | <b>0,50</b>            | <b>0.0117</b> |
|   |                              | <b>0,25</b>            | <b>0.0065</b> |
|   |                              | <b>Slope value</b>     | <b>0,02</b>   |
| <b>Important notice</b>   |                              |                        |               |
| <p>For Yumizen G400 DDi and Yumizen G200, make sure that the end time is set to 180 seconds in the D-Dimer setup and the curve type <b>Lin-Lin p-p</b> is selected in <b>Calibration &gt; D-Dimer &gt; ug FEU/ml</b>.<br/>         For more information, please contact your local HORIBA Medical representative.</p> |                              |                        |               |