

ABX Minocal



LOT

CX 495

CAL



(Exp.) 2024-07-05
(YYYY - MM - DD)

Rev 1

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|-------------|---------|--------|--------|----------------------|--------|--------|---------------------|--------|-------------------------|
| | | MICROS | MICROS | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | |
| | | EMI CRP | LC-767G | MS CRP | XLR | DX NEXUS DF NEXUS | MS60 | | 60 60 C+ ES60 | XL80 | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.1 | 8.9 | 9.2 | 9.0 | 9.3 | 9.1 | | 9.0 | 9.0 | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.43 | 4.45 | 4.42 | 4.47 | 4.47 | 4.40 | | 4.49 | 4.47 | ± 0.06 |
| HB HGB | g/dl | 13.1 | 13.2 | 13.1 | 13.3 | 13.2 | 13.1 | | 13.3 | 13.3 | ± 0.2 |
| | g/l | 131 | 132 | 131 | 133 | 132 | 131 | | 133 | 133 | ± 2 |
| | mmol/l | 8.14 | 8.20 | 8.14 | 8.26 | 8.20 | 8.14 | | 8.26 | 8.26 | ± 0.12 |
| HT HCT | % | 37.0 | 37.2 | 35.8 | 36.2 | 36.0 | 35.4 | | 36.4 | 36.2 | ± 1.0 |
| | l/l | 0.370 | 0.372 | 0.358 | 0.362 | 0.360 | 0.354 | | 0.364 | 0.362 | ± 0.010 |
| PLA PLT | 10 ³ /mm ³ ;10 ⁹ /l | 231 | 230 | 221 | 231 | 234 | 231 | | 233 | 231 | ± 10 |
| VMP MPV | µm ³ ; fl | 8.1 | 8.4 | N/A | N/A | N/A | N/A | | N/A | N/A | ± 0.5 |

Ref: TEMP-0387 Rev.51 BACK / VERSO 1300090308