

# ABX Minocal



**LOT**

**CX 491**

**CAL**



**(Exp.) 2024-03-05**

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	WHITEDIFF								TOLERANCES TOLERANCE
		YUMIZEN V1.0 to V2.x				YUMIZEN since V3				
		H550	H500 OT H500 CT			H500 OT	H500 CT H550			
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>8.75</b>	<b>8.75</b>			<b>8.70</b>	<b>8.70</b>			± 0.20
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>4.57</b>	<b>4.57</b>			<b>4.57</b>	<b>4.57</b>			± 0.06
HB HGB	g/dl	<b>12.8</b>	<b>12.8</b>			<b>12.9</b>	<b>12.9</b>			± 0.2
	g/l	<b>128</b>	<b>128</b>			<b>129</b>	<b>129</b>			± 2
	mmol/l	<b>7.95</b>	<b>7.95</b>			<b>8.01</b>	<b>8.01</b>			± 0.12
HT HCT	%	<b>37.5</b>	<b>37.5</b>			<b>36.7</b>	<b>36.3</b>			± 1.0
	l/l	<b>0.375</b>	<b>0.375</b>			<b>0.367</b>	<b>0.363</b>			± 0.010
PLA PLT	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>248</b>	<b>248</b>			<b>261</b>	<b>268</b>			± 10
VMP MPV	μm <sup>3</sup> ; fl	<b>10.6</b>	<b>10.6</b>			N/A	N/A			± 0.5

Ref: TEMP-0387 Rev.51 BACK / VERSO 1300113555