

ABX Difftrol



LOT PX 436
Rev 1

CONTROL

(Exp.) 2022-09-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Whitediff														
		CONTROL				L	CONTROL				N	CONTROL				H
		YUMIZEN				TOLERANCES TOLERANCE	YUMIZEN				TOLERANCES TOLERANCE	YUMIZEN				TOLERANCES TOLERANCE
		H550 V1.0 to V2.x	H500 OT H500 CT	H500 OT Since V3	H500 CT H550		H550 V1.0 to V2.x	H500 OT H500 CT	H500 OT Since V3	H500 CT H550		H550 V1.0 to V2.x	H500 OT H500 CT	H500 OT Since V3	H500 CT H550	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.90	2.90	3.00	3.00	± 0.40	8.45	8.45	8.55	8.54	± 1.00	18.40	18.40	18.37	18.58	± 2.20
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.38	2.38	2.43	2.38	± 0.16	4.73	4.73	4.68	4.63	± 0.20	5.24	5.24	5.08	5.04	± 0.25
HB HGB	g/dl	6.2	6.2	6.1	6.3	± 0.4	13.3	13.3	13.3	13.3	± 0.5	15.8	15.8	15.8	15.7	± 0.6
	g/l	62	62	61	63	± 4	133	133	133	133	± 5	158	158	158	157	± 6
	mmol/l	3.85	3.85	3.79	3.91	± 0.25	8.26	8.26	8.26	8.26	± 0.31	9.81	9.81	9.81	9.75	± 0.37
HT HCT	%	18.8	18.8	17.0	16.6	± 1.5	40.9	40.9	36.0	35.9	± 2.0	48.2	48.2	42.1	42.2	± 2.5
	l/l	0.188	0.188	0.170	0.166	± 0.015	0.409	0.409	0.360	0.359	± 0.020	0.482	0.482	0.421	0.422	± 0.025
VGM MCV	µm ³ ; fl	79.0	79.0	69.9	69.9	± 5.0	85.0	85.0	77.0	77.5	± 5.0	92.0	92.0	82.9	83.8	± 5.0
TGMH MCH	pg	26.1	26.1	25.1	26.5	± 2.0	28.1	28.1	28.4	28.7	± 2.0	30.2	30.2	31.1	31.2	± 2.5
	fmol	1.62	1.62	1.56	1.65	± 0.12	1.75	1.75	1.76	1.78	± 0.12	1.88	1.88	1.93	1.94	± 0.16
CCMH MCHC	g/dl	33.0	33.0	35.9	37.9	± 3.0	33.1	33.1	36.9	37.1	± 3.0	32.8	32.8	37.5	37.2	± 3.0
	g/l	330	330	359	379	± 30	331	331	369	371	± 30	328	328	375	372	± 30
	mmol/l	20.49	20.49	22.29	23.54	± 1.86	20.56	20.56	22.91	23.04	± 1.86	20.37	20.37	23.29	23.10	± 1.86
IDR-SD RDW-SD	fl	43.0	43.0	42.8	36.8	± 8.0	39.5	39.5	41.3	35.9	± 8.0	41.0	41.0	43.6	37.4	± 8.0
IDR-CV RDW-CV	%	17.0	17.0	14.7	15.4	± 4.0	14.5	14.5	12.7	13.4	± 4.0	14.0	14.0	12.3	12.8	± 4.0
PLA PLT	10 ⁹ /mm ³ ; 10 ⁹ /l	77	77	86	83	± 20	268	268	259	246	± 30	520	520	502	493	± 50
VMP MPV	µm ³ ; fl	10.1	10.1	10.6	11.0	± 2.0	13.2	13.2	11.3	11.4	± 2.0	11.9	11.9	10.7	10.9	± 2.0
NEU	10 ⁹ /mm ³ ; 10 ⁹ /l	1.22	1.22	1.24	1.30	± 0.35	3.99	3.99	3.94	4.01	± 0.90	12.79	12.79	12.44	12.84	± 1.90
	%	42.1	42.1	41.2	43.2	± 10.0	47.2	47.2	46.1	46.9	± 10.0	69.5	69.5	67.7	69.1	± 10.0
LYM	10 ⁹ /mm ³ ; 10 ⁹ /l	1.11	1.11	1.19	1.15	± 0.33	3.63	3.63	3.71	3.70	± 0.70	3.75	3.75	3.78	3.73	± 1.50
	%	38.4	38.4	39.8	38.3	± 12.0	42.9	42.9	43.4	43.3	± 8.0	20.4	20.4	20.6	20.1	± 8.0
MON	10 ⁹ /mm ³ ; 10 ⁹ /l	0.26	0.26	0.27	0.27	± 0.26	0.41	0.41	0.41	0.41	± 0.41	0.64	0.64	0.70	0.72	± 0.64
	%	9.0	9.0	8.9	8.9	± 8.9	4.9	4.9	4.8	4.8	± 4.8	3.5	3.5	3.8	3.9	± 3.5
EOS	10 ⁹ /mm ³ ; 10 ⁹ /l	0.18	0.18	0.17	0.17	± 0.17	0.28	0.28	0.29	0.27	± 0.27	0.85	0.85	0.88	0.87	± 0.85
	%	6.3	6.3	5.7	5.6	± 5.6	3.3	3.3	3.4	3.2	± 3.2	4.6	4.6	4.8	4.7	± 4.6
BAS	10 ⁹ /mm ³ ; 10 ⁹ /l	0.12	0.12	0.13	0.12	± 0.12	0.14	0.14	0.20	0.15	± 0.14	0.37	0.37	0.57	0.41	± 0.37
	%	4.2	4.2	4.4	4.0	± 4.0	1.7	1.7	2.3	1.8	± 1.7	2.0	2.0	3.1	2.2	± 2.0
IMG	10 ⁹ /mm ³ ; 10 ⁹ /l	N/A	N/A	0.09	0.11	± 0.09	N/A	N/A	0.33	0.28	± 0.28	N/A	N/A	1.03	0.91	± 0.91
	%	N/A	N/A	2.9	3.5	± 2.9	N/A	N/A	3.9	3.3	± 3.3	N/A	N/A	5.6	4.9	± 4.9

1300105209
FRONT / RECTO
Ref: TEMP-0821 Rev.46