



Certificate of Traceability

Manufacturer Certification

This certifies that **Horiba Medical**, using **Pointe Hemoglobin A1c – Pentra C200** has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **European Reference Laboratory ESRL#13**

The system evaluated was:

Instrument: Pentra C200	Calibrator Lot: 211501	Application: $Y=(1.15 \times X)-0.68$
Reagent Lot: 322201	Calibrator Assigned Values: 0%, 5%, 7.9%, 11.8%, 15.2%	

Date of Certification: June 1, 2024

Certification Expires: June 1, 2025

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/supervisor