

**HORIBA ABX SAS**  
HORIBA ABX SAS  
Parc Euromédecine  
Rue du Caducée - BP 7290  
34184 MONTPELLIER Cedex 4  
FRANCE  
Tél.: 33 (0) 4 67 14 15 16  
Fax.: 33 (0) 4 67 14 15 17

**LOT**

**MC 432**

**CONTROL**



**(Exp.) 2024-10-05**  
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio													
				CONTROL			1	CONTROL			2	CONTROL			3	TOLERANCES TOLERANCE	
				MICROS		PENTRA	TOLERANCES TOLERANCE	MICROS		PENTRA	TOLERANCES TOLERANCE	MICROS		PENTRA			
				EMI CRP	LC-767G	MS CRP		EMI CRP	LC-767G	MS CRP		EMI CRP	LC-767G	MS CRP			
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.0</b>	<b>2.0</b>	<b>2.1</b>	± 0.4	<b>7.7</b>		<b>7.6</b>	<b>7.8</b>	± 1.0		<b>20.4</b>	<b>20.4</b>	<b>21.0</b>	± 2.2		
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.38</b>	<b>2.37</b>	<b>2.33</b>	± 0.15	<b>4.49</b>	<b>4.46</b>	<b>4.49</b>	± 0.18	<b>5.50</b>	<b>5.47</b>	<b>5.54</b>	± 0.25				
HB HGB	g/dl	<b>6.3</b>	<b>6.3</b>	<b>6.0</b>	± 0.4	<b>13.5</b>	<b>13.3</b>	<b>13.4</b>	± 0.5	<b>17.5</b>	<b>17.4</b>	<b>17.7</b>	± 0.7				
	g/l	<b>63</b>	<b>63</b>	<b>60</b>	± 4	<b>135</b>	<b>133</b>	<b>134</b>	± 5	<b>175</b>	<b>174</b>	<b>177</b>	± 7				
	mmol/l	<b>3.91</b>	<b>3.91</b>	<b>3.73</b>	± 0.25	<b>8.38</b>	<b>8.26</b>	<b>8.32</b>	± 0.31	<b>10.87</b>	<b>10.81</b>	<b>10.99</b>	± 0.44				
HT HCT	%	<b>17.1</b>	<b>17.1</b>	<b>16.5</b>	± 2.0	<b>37.3</b>	<b>36.6</b>	<b>36.6</b>	± 2.5	<b>48.4</b>	<b>47.6</b>	<b>47.9</b>	± 3.0				
	l/l	<b>0.171</b>	<b>0.171</b>	<b>0.165</b>	± 0.020	<b>0.373</b>	<b>0.366</b>	<b>0.366</b>	± 0.025	<b>0.484</b>	<b>0.476</b>	<b>0.479</b>	± 0.030				
VGM MCV	µm <sup>3</sup> ; fl	<b>72</b>	<b>72</b>	<b>71.0</b>	± 4	<b>83</b>	<b>82</b>	<b>81.5</b>	± 4	<b>88</b>	<b>87</b>	<b>86.5</b>	± 4				
TGMH MCH	pg	<b>26.5</b>	<b>26.6</b>	<b>25.8</b>	± 2.4	<b>30.1</b>	<b>29.8</b>	<b>29.8</b>	± 2.4	<b>31.8</b>	<b>31.8</b>	<b>31.9</b>	± 2.8				
	fmol	<b>1.64</b>	<b>1.65</b>	<b>1.60</b>	± 0.15	<b>1.87</b>	<b>1.85</b>	<b>1.85</b>	± 0.15	<b>1.98</b>	<b>1.98</b>	<b>1.98</b>	± 0.18				
CCMH MCHC	g/dl	<b>36.8</b>	<b>36.9</b>	<b>36.3</b>	± 3.0	<b>36.2</b>	<b>36.4</b>	<b>36.6</b>	± 3.0	<b>36.2</b>	<b>36.6</b>	<b>36.9</b>	± 3.0				
	g/l	<b>368</b>	<b>369</b>	<b>363</b>	± 30	<b>362</b>	<b>364</b>	<b>366</b>	± 30	<b>362</b>	<b>366</b>	<b>369</b>	± 30				
	mmol/l	<b>22.83</b>	<b>22.93</b>	<b>22.52</b>	± 1.86	<b>22.50</b>	<b>22.58</b>	<b>22.74</b>	± 1.86	<b>22.45</b>	<b>22.71</b>	<b>22.94</b>	± 1.86				
IDR RDW	%	<b>14.5</b>	<b>14.0</b>	<b>14.5</b>	± 3.0	<b>14.0</b>	<b>13.0</b>	<b>13.0</b>	± 3.0	<b>13.5</b>	<b>12.5</b>	<b>11.5</b>	± 3.0				
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>66</b>	<b>73</b>	<b>63</b>	± 20	<b>245</b>	<b>243</b>	<b>240</b>	± 40	<b>468</b>	<b>477</b>	<b>507</b>	± 60				
VPM MPV	µm <sup>3</sup> ; fl	<b>7.3</b>	<b>7.9</b>	<b>8.7</b>	± 2.0	<b>7.9</b>	<b>8.0</b>	<b>8.6</b>	± 2.0	<b>7.3</b>	<b>7.4</b>	<b>7.7</b>	± 2.0				
LYMPHO	#	<b>1.3</b>	<b>1.3</b>	N/A	± 0.2	<b>2.6</b>	<b>2.5</b>	N/A	± 0.5	<b>3.5</b>	<b>3.4</b>	N/A	± 1.1				
	%	<b>66.0</b>	<b>65.5</b>	N/A	± 8.0	<b>33.5</b>	<b>33.0</b>	N/A	± 6.0	<b>17.0</b>	<b>16.5</b>	N/A	± 5.0				
MONO	#	<b>0.2</b>	<b>0.2</b>	N/A	± 0.2	<b>0.5</b>	<b>0.5</b>	N/A	± 0.5	<b>1.0</b>	<b>1.1</b>	N/A	± 1.0				
	%	<b>8.0</b>	<b>8.0</b>	N/A	± 7.2	<b>6.5</b>	<b>7.0</b>	N/A	± 6.3	<b>5.0</b>	<b>5.5</b>	N/A	± 5.0				
GRANULO	#	<b>0.5</b>	<b>0.5</b>	N/A	± 0.2	<b>4.6</b>	<b>4.6</b>	N/A	± 0.8	<b>15.9</b>	<b>15.9</b>	N/A	± 2.0				
	%	<b>26.0</b>	<b>26.5</b>	N/A	± 10.0	<b>60.0</b>	<b>60.0</b>	N/A	± 10.0	<b>78.0</b>	<b>78.0</b>	N/A	± 10.0				
CRP	mg/dl	<b>0.49</b>	<b>0.44</b>	<b>0.44</b>	± 0.40	<b>2.78</b>	<b>2.86</b>	<b>2.64</b>	± 0.80	<b>5.01</b>	<b>4.87</b>	<b>5.10</b>	± 1.50				
	mg/l	<b>4.9</b>	<b>4.4</b>	<b>4.4</b>	± 4.0	<b>27.8</b>	<b>28.6</b>	<b>26.4</b>	± 8.0	<b>50.1</b>	<b>48.7</b>	<b>51.0</b>	± 15.0				

Ref: TEMP-1070 Rev.43 BACKVERSO 1300092170