

# Yumizen G PT 5 for Yumizen G Line



<b>Reference No.:</b>		<b>1300036338</b>		
<b>Lot No.:</b>		<b>940114 (rev. 1)</b>		
<b>Expiry date:</b>		<b>2026-01-31</b>		
<b>Coagulometers</b>		<b>ISI</b>	<b>MASTER CURVE</b>	
<b>Optical</b>	<b>Yumizen G Line</b>		<b>%</b>	<b>sec</b>
		<b>100</b>	<b>13,3</b>	
		<b>50</b>	<b>18,9</b>	
		<b>33</b>	<b>24,6</b>	
		<b>25</b>	<b>30,0</b>	
		<b>1,11</b>	<b>MNPT 13,3</b>	
<b>Important notice</b>				
<ul style="list-style-type: none"> <li>- The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</li> <li>- The efficiency of the reconstitution of the reagent may be improved by performing it at 37°C.</li> </ul>				