

ABX Minocal



LOT

CX 497
Rev 1

CAL

(Exp.) **2024-09-05**
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|--------------|--------------|--------------|--------------|----------------------|--------------|--------|---------------------|--------------|-------------------------|
| | | MICROS | MICROS | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | |
| | | EMI CRP | LC-767G | MS CRP | XLR | DX NEXUS DF NEXUS | MS60 | | 60 60 C+ ES60 | XL80 | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.2 | 9.0 | 9.3 | 9.1 | 9.3 | 9.1 | | 9.2 | 9.1 | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.47 | 4.44 | 4.44 | 4.47 | 4.51 | 4.45 | | 4.50 | 4.47 | ± 0.06 |
| HB HGB | g/dl | 13.0 | 12.9 | 13.1 | 13.2 | 13.1 | 13.0 | | 13.1 | 13.2 | ± 0.2 |
| | g/l | 130 | 129 | 131 | 132 | 131 | 130 | | 131 | 132 | ± 2 |
| | mmol/l | 8.07 | 8.01 | 8.14 | 8.20 | 8.14 | 8.07 | | 8.14 | 8.20 | ± 0.12 |
| HT HCT | % | 37.1 | 36.4 | 35.7 | 36.0 | 36.1 | 35.6 | | 36.2 | 36.0 | ± 1.0 |
| | l/l | 0.371 | 0.364 | 0.357 | 0.360 | 0.361 | 0.356 | | 0.362 | 0.360 | ± 0.010 |
| PLA PLT | 10 ³ /mm ³ ;10 ⁹ /l | 234 | 239 | 229 | 227 | 237 | 226 | | 234 | 227 | ± 10 |
| VMP MPV | µm ³ ; fl | 9.0 | 9.1 | N/A | N/A | N/A | N/A | | N/A | N/A | ± 0.5 |

Ref: TEMP-0387 Rev.52 BACK / VERSO 1300090308