

ABX Minocal



LOT

CX 498
Rev 1

CAL

(Exp.) **2024-10-05**
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP	LC-767G	MS CRP	XLR	DX NEXUS DF NEXUS	MS60		60 60 C+ ES60	XL80	
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.1	9.0	9.2	9.1	9.2	9.0		8.9	9.1	± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.38	4.38	4.34	4.42	4.44	4.38		4.45	4.42	± 0.06
HB HGB	g/dl	12.7	12.7	12.7	12.9	12.8	12.8		12.9	12.9	± 0.2
	g/l	127	127	127	129	128	128		129	129	± 2
	mmol/l	7.89	7.89	7.89	8.01	7.95	7.95		8.01	8.01	± 0.12
HT HCT	%	35.7	35.9	34.9	35.4	35.1	34.8		35.6	35.4	± 1.0
	l/l	0.357	0.359	0.349	0.354	0.351	0.348		0.356	0.354	± 0.010
PLA PLT	10 ³ /mm ³ ;10 ⁹ /l	229	228	216	225	236	224		231	225	± 10
VMP MPV	µm ³ ; fl	8.6	8.8	N/A	N/A	N/A	N/A		N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.52 BACK / VERSO 1300090308