



Manufactured for
HORIBA Instruments Incorporated:
Pointe Brand
5449 Research Drive
Canton, MI 48188



LOT 401505

SAP 5390015525



REF T7535-CTL



2024-12-31

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	Unité Unit Einheit Unidad Unità Unidade Enhed Enhet Μονάδα Jednostka Enhet	Level I Exp. Date: 2024-12-31 Valeur titrée Assigned value Deklarierter Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tilddelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd			Level II Exp. Date: 2024-12-31 Valeur titrée Assigned value Deklarierter Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tilddelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd		
		Mean	SD	Range	Mean	SD	Range
		Transferrin*	mg/dL	212	21	170-254	521
	g/L	2.12	0.21	1.70-2.54	5.21	0.26	4.69-5.73

*Values are standardized to the CAP/BCR/IFCC Reference Prep. for Proteins in Human Serum (RPPHS), lot 91/0619