



# Yumizen G PT Reco 10 for Yumizen G Line



<b>Reference No.:</b>		<b>1300036376</b>		
<b>Lot No.:</b>		<b>940115 (rev. 1)</b>		
<b>Expiry date:</b>		<b>2026-01-31</b>		
<b>Coagulometers</b>		<b>ISI</b>	<b>MASTER CURVE</b>	
<b>Optical</b>	<b>Yumizen G Line</b>		<b>%</b>	<b>SEC</b>
		<b>100</b>	<b>10.7</b>	
		<b>50</b>	<b>14.2</b>	
		<b>33</b>	<b>17.9</b>	
		<b>25</b>	<b>21.3</b>	
		<b>1,11</b>	<b>MNPT</b>	<b>10.7</b>
<b>Important notice</b>				
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>For Yumizen G200, Yumizen G400 and Yumizen G400 DDi, the <b>Min. Time</b> must be set to 7 seconds and the <b>Lag time</b> to 5 seconds in the PT setup. Please refer to the Yumizen G PT Reco 10 reagent notice.</p>				