



Yumizen G CAL ***for Yumizen G Line***

Reference No.:		1300036416
Lot No.:		941207 (rev. 1)
Expiry date:		2026-12-31
Parameter	Unit	Value
Prothrombin	%	90
Fibrinogen	g/L	2,86
	mg/dL	286
Antithrombin	%	93
	IU/mL	0,93
Important notice		
Dilution serie must be prepared. Please refer to the notice.		