



Yumizen G PT 5 for Yumizen G Line



Reference No.:		1300036338		
Lot No.:		941015 (rev. 1)		
Expiry date:		2026-10-31		
Coagulometers		ISI	MASTER CURVE	
Optical	Yumizen G Line		1,24	%
		100		12,7
		50		17,9
		33		23,3
		25		28,3
			MNPT	12,7
Important notice				
<ul style="list-style-type: none"> - The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT. - The efficiency of the reconstitution of the reagent may be improved by performing it at 37°C. 				