



Yumizen G PT Reco 10 for Yumizen G Line



Reference No.:		1300036376		
Lot No.:		940604 (rev. 1)		
Expiry date:		2026-06-30		
Coagulometers		ISI	MASTER CURVE	
Optical	Yumizen G Line		%	SEC
		100	10.7	
		50	15.0	
		33	19.5	
		25	23.7	
		1,01	MNPT	10.7
Important notice				
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>For Yumizen G200, Yumizen G400 and Yumizen G400 DDi, the Min. Time must be set to 7 seconds and the Lag time to 5 seconds in the PT setup. Please refer to the Yumizen G PT Reco 10 reagent notice.</p>				