

# ABX Minocal



**LOT**

**CX 506**

**CAL**



**(Exp.) 2025-06-05**  
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP	LC-767G	MS CRP	XLR	DX NEXUS DF NEXUS	MS60		60 60 C+ ES60	XL80	
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>9.1</b>	<b>9.0</b>	<b>9.3</b>	<b>9.1</b>	<b>9.3</b>	<b>9.2</b>		<b>9.3</b>	<b>9.1</b>	± 0.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>4.37</b>	<b>4.37</b>	<b>4.34</b>	<b>4.36</b>	<b>4.40</b>	<b>4.34</b>		<b>4.39</b>	<b>4.36</b>	± 0.06
HB HGB	g/dl	<b>12.9</b>	<b>13.0</b>	<b>13.0</b>	<b>13.2</b>	<b>13.0</b>	<b>12.9</b>		<b>13.2</b>	<b>13.2</b>	± 0.2
	g/l	<b>129</b>	<b>130</b>	<b>130</b>	<b>132</b>	<b>130</b>	<b>129</b>		<b>132</b>	<b>132</b>	± 2
	mmol/l	<b>8.01</b>	<b>8.07</b>	<b>8.07</b>	<b>8.20</b>	<b>8.07</b>	<b>8.01</b>		<b>8.20</b>	<b>8.20</b>	± 0.12
HT HCT	%	<b>36.5</b>	<b>36.7</b>	<b>35.2</b>	<b>36.0</b>	<b>35.9</b>	<b>35.4</b>		<b>36.2</b>	<b>36.0</b>	± 1.0
	l/l	<b>0.365</b>	<b>0.367</b>	<b>0.352</b>	<b>0.360</b>	<b>0.359</b>	<b>0.354</b>		<b>0.362</b>	<b>0.360</b>	± 0.010
PLA PLT	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>229</b>	<b>231</b>	<b>224</b>	<b>228</b>	<b>236</b>	<b>220</b>		<b>228</b>	<b>228</b>	± 10
VMP MPV	µm <sup>3</sup> ; fl	<b>8.1</b>	<b>8.1</b>	N/A	N/A	N/A	N/A		N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.52 BACK / VERSO 1300090308