



Manufactured for  
HORIBA Instruments Incorporated:  
Pointe Brand  
5449 Research Drive  
Canton, MI 48188



**LOT** 421402

**SAP** 5390015523



**REF** F7505-CTL



2026-05-31

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	Unité Unit Einheit Unidad Unità Unidade Enhed Enhet Μονάδα Jednostka Enhet	Level I Exp. Date: 2026-05-31 Valeur titrée Assigned value Deklarierter Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tilddelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd			Level II Exp. Date: 2026-05-31 Valeur titrée Assigned value Deklarierter Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tilddelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd		
		Mean	SD	Range	Mean	SD	Range
<b>Ferritin</b> (1st I.S. Standardization)	ng/mL	103	7.5	88-118	403	30.0	343-463
<b>Ferritin</b> (3rd I.S. Standardization)	ng/mL	123	9.0	105-141	480	36.0	408-552

Ferritin values are provided with two options: 1<sup>st</sup> or 3<sup>rd</sup> I.S. (International Standard) Values.

It is recommended to use the same standardized system as was chosen when selecting values for the Pointe Ferritin Calibrators (F7505-CAL).