



**HORIBA ABX SAS**  
 Parc Euromédecine  
 Rue du Caducée  
 B.P. 7290  
 34184 Montpellier cedex 4  
 FRANCE  
 www.horiba-abx.com



A03A00009B

# Yumizen G PT Liq 4 for Yumizen G Line



<b>Reference No.:</b>		<b>1300036373</b>		
<b>Lot No.:</b>		<b>950108 (rev. 1)</b>		
<b>Expiry date:</b>		<b>2026-07-31</b>		
<b>Coagulometers</b>		<b>ISI</b>	<b>MASTER CURVE</b>	
<b>Optical</b>	<b>Yumizen G Line</b>		<b>1,20</b>	<b>%</b>
		<b>100</b>		<b>12,2</b>
		<b>50</b>		<b>18,3</b>
		<b>33</b>		<b>24,5</b>
		<b>25</b>		<b>30,4</b>
	<b>MNPT</b>	<b>12,2</b>		
<b>Important notice</b>				
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p>				