



HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée
 B.P. 7290
 34184 Montpellier cedex 4
 FRANCE
 www.horiba-abx.com



A03A00009B

Yumizen G PT Liq 4 for Yumizen G Line



Reference No.:		1300036373		
Lot No.:		950208 (rev. 1)		
Expiry date:		2026-08-31		
Coagulometers		ISI	MASTER CURVE	
Optical	Yumizen G Line		1,27	%
		100		11,7
		50		17,0
		33		22,4
		25		27,5
	MNPT	11,7		
Important notice				
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p>				