

# Yumizen G PT Liq 4 for Yumizen G Line



|  |                       |                        |                     |             |
|--|-----------------------|------------------------|---------------------|-------------|
| <b>Reference No.:</b>  |                       | <b>1300036373</b>      |                     |             |
| <b>Lot No.:</b>  |                       | <b>950219 (rev. 1)</b> |                     |             |
| <b>Expiry date:</b>  |                       | <b>2026-08-31</b>      |                     |             |
| <b>Coagulometers</b>   |                       | <b>ISI</b>             | <b>MASTER CURVE</b> |             |
| <b>Optical</b>   | <b>Yumizen G Line</b> |                        | <b>1,29</b>         | <b>%</b>    |
|  |                       | <b>100</b>             |                     | <b>11,8</b> |
|  |                       | <b>50</b>              |                     | <b>17,1</b> |
|  |                       | <b>33</b>              |                     | <b>22,5</b> |
|  |                       | <b>25</b>              |                     | <b>27,6</b> |
|  | <b>MNPT</b>           | <b>11,8</b>            |                     |             |
| <b>Important notice</b>  |                       |                        |                     |             |
| <p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> |                       |                        |                     |             |