

LOT
CX 508
CAL

(Exp.) 2025-08-05
 (YYYY - MM - DD)

 Rev **1**

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP	LC-767G	MS CRP	XLR	DX NEXUS DF NEXUS	MS60		60 60 C+ ES60	XL80	
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.1	9.0	9.3	9.0	9.3	8.9		8.9	9.0	± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.57	4.55	4.50	4.55	4.61	4.53		4.61	4.55	± 0.06
HB HGB	g/dl	13.6	13.6	13.7	13.8	13.7	13.6		13.8	13.8	± 0.2
	g/l	136	136	137	138	137	136		138	138	± 2
	mmol/l	8.45	8.45	8.51	8.57	8.51	8.45		8.57	8.57	± 0.12
HT HCT	%	38.2	38.0	36.9	37.5	37.1	36.7		37.8	37.5	± 1.0
	l/l	0.382	0.380	0.369	0.375	0.371	0.367		0.378	0.375	± 0.010
PLA PLT	10 ³ /mm ³ ;10 ⁹ /l	230	233	227	229	231	225		230	229	± 10
VMP MPV	µm ³ ; fl	7.9	8.1	N/A	N/A	N/A	N/A		N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.52 BACK / VERSO 1300090308