



Manufactured for
HORIBA Instruments Incorporated:
Pointe Brand
5449 Research Drive
Canton, MI 48188



LOT 430903

SAP 5390013053

REF M7562-CTL



2027-07-31

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	Unité Unit Einheit Unidad Unità Unidade Enhed Enhet Μονάδα Jednostka Enhet	Level I Exp. Date: 2027-07-31 Valeur titrée Assigned value Deklariertes Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tilldelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd			Level II Exp. Date: 2027-07-31 Valeur titrée Assigned value Deklariertes Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tilldelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd		
		Mean	SD	Range	Mean	SD	Range
Microalbumin	mg/dL	1.3	0.2	0.9 – 1.7	6.4	0.95	4.5 – 8.3
	mg/L	13	2	9 - 17	64	9.5	45 – 83

*Mean and range calculated using data from Roche Hitachi and Mindray BS200 analyzers.