


## EU Declaration of Conformity

(N° dc90008ben)

### WE THE MANUFACTURER

Name	<b>HORIBA ABX SAS</b>
 Address	Parc Euromédecine Rue du Caducée BP 7290 34184 Montpellier Cedex 4 FRANCE
Single Registration Number	FR-MF-000000320

### TAKE SOLE RESPONSIBILITY FOR AND HEREBY DECLARE THAT THE PRODUCT(S)

Device category	<b>Hematology Calibrator</b>
Product name	<b>ABX Minocal</b>
Models	<b>2032002</b>
Basic UDI-DI	<b>361023minocalJC</b>
Country of origin	<b>USA</b>

### Intended Use

**ABX Minocal** is a multiparameter blood calibrator intended for *in vitro* diagnostic use and designed for use in calibration of hematology blood cell counters for the following CBC parameters (WBC, RBC, HGB, HCT, MCV, PLT, MPV) in clinical laboratories.

Parameters can be different according to the instrument, please refer to the assay value data sheet for specific instrument models.

## MEET(S) THE PROVISIONS OF THE FOLLOWING DIRECTIVES, REGULATIONS, STANDARDS AND COMMON SPECIFICATIONS

Regulations	Regulation (EU) 2017/746 on <i>in vitro</i> diagnostic medical devices <b>Risk Class:</b> A <input type="checkbox"/> B <input checked="" type="checkbox"/> C <input type="checkbox"/> D <input type="checkbox"/>	
IVDR conformity assessment route	<input checked="" type="checkbox"/> ANNEX IX (Chap I & III, chap II section 4) + ANNEX IV ( <i>Class B &amp; C devices excluding self-testing and near patient testing devices</i> )	EU CERTIFICATE #: IVDR 745367 Name of Notified Body: BSI Group The Netherlands B.V. Notified Body Identification: 2797
Common Specifications	Not applicable	

Montpellier, France  
2025/07/23

**Claire MALLIÉ**  
Quality & Regulatory Affairs Junior  
Director / PRRC

