



HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée
 B.P. 7290
 34184 Montpellier Cedex 4
 FRANCE
 www.horiba.com



LOT 38020 (rev01)

REF : 1300071247

SAP 1300082994



2027/05/16

(FR) Constituant	Yumizen C1200	Unité	Valeur titrée	Intervalle de confiance	
(EN) Constituent		Unit	Assigned value	Confidence range	
(DE) Bestandteil		Einheit	Deklariertes Wert	Vertrauensintervalle	
(ES) Componente		Unidad	Valor asignado	Intervalos de confianza	
(IT) Componente		Unità	Valori assegnati	Livelli di fiducia	
(PT) Componente		Unidade	Valor determinado	Intervalo de confiança	
(DA) Bestanddel		Enhed	Tildelt værdi	Konfidensområde	
(SV) Konstituent		Enhet	Tildelat värde	Konfidensintervall	
(EL) συστατικό		Μονάδα	Αναγραφόμενη τιμή	Εύρος εμπιστοσύνης	
(PL) Składnik		Jednostka	Wartość przypisana	Zakres ufności	
(NO) Bestanddel	Enhet	Tildelt verd	Konfidensspekter		
C		U	\oplus	Confidence range	1 SD
Yumizen C1200 CRP High sensitive method (1300023877)	•	mg/L	1.00	0.80 - 1.20	0.10