



Yumizen G DDi 2 for Yumizen G15X0x and Yumizen G405



D 4 9 5 0 8 0 1 3 6 3 9 1 0 8 2 6 D



D 5 4 0 0 2 6 1 0 3 0 0 1 9 1 0 2 D



D 6 0 0 1 1 9 0 1 0 0 0 5 5 0 0 5 D



D 7 0 0 2 7 0 0 2 5 0 1 4 0 1 6 0 D

Reference No.:		1300036391	
Lot No.:		950801 (rev. 1)	
Expiry date:		2026-08-31	
Coagulometers		MASTER CURVE	
Optical	Yumizen G15X0x/G405	µg(FEU)/mL	ΔOD
		4.00	0.2610
		3.00	0.1910
		2.00	0.1190
		1.00	0.0550
		0.50	0.0270
		0.25	0.0140
Slope value		0,07	

Yumizen G DDi 2 for Yumizen G8X0x



D 4 9 5 0 8 0 1 3 6 3 9 1 0 8 2 6 D



D 5 4 0 0 2 5 1 5 3 0 0 1 8 9 0 2 D



D 6 0 0 1 2 1 5 1 0 0 0 5 6 5 0 5 D



D 7 0 0 2 6 5 0 2 5 0 1 2 0 1 7 7 D

Reference No.:		1300036391	
Lot No.:		950801 (rev. 1)	
Expiry date:		2026-08-31	
Coagulometers		MASTER CURVE	
Optical	Yumizen G8X0x	µg(FEU)/mL	ΔOD
		4.00	0.2515
		3.00	0.1890
		2.00	0.1215
		1.00	0.0565
		0.50	0.0265
		0.25	0.0120
Slope value		0,05	



Yumizen G DDi 2 for Yumizen G400 DDi/G200

Reference No.:		1300036391	
Lot No.:		950801 (rev. 1)	
Expiry date:		2026-08-31	
Coagulometers		MASTER CURVE	
		µg(FEU)/mL	ΔOD
Optical	Yumizen G400 DDi/G200	4.00	0.1008
		3.00	0.0730
		2.00	0.0480
		1.00	0.0238
		0.50	0.0115
		0.25	0.0054
		Slope value	0,02
Important notice			
<p>For Yumizen G400 DDi and Yumizen G200, make sure that the end time is set to 180 seconds in the D-Dimer setup and the curve type Lin-Lin p-p is selected in Calibration > D-Dimer > ug FEU/ml. For more information, please contact your local HORIBA Medical representative.</p>			