



Manufactured for
HORIBA Instruments Incorporated:
Pointe Brand
5449 Research Drive
Canton, MI 48188



LOT 522405

SAP 5390015525

REF T7535-CTL



2027-01-31

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	Unité Unit Einheit Unidad Unità Unidade Enhed Enhet Μονάδα Jednostka Enhet	Level I Exp. Date: 2027-01-31 Valeur titrée Assigned value Deklariertes Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tilldelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd			Level II Exp. Date: 2027-01-31 Valeur titrée Assigned value Deklariertes Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tilldelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd		
		Mean	SD	Range	Mean	SD	Range
		Transferrin*	mg/dL	245	24.5	196-294	483
	g/L	2.45	0.245	1.96-2.94	4.83	0.24	4.35-5.31

*Values are standardized to the CAP/BCR/IFCC Reference Prep. for Proteins in Human Serum (RPPHS), lot 91/0619