

ABX Pentra Etching CP

REF A11A01769

REAGENT 25 mL

IVD CE Rx Only

HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée
BP 7290
34184 Montpellier Cedex 4
FRANCE



- Pentra C200
- Pentra C400
- ABX Pentra 400

Cleaning solution used on the ISE module of the HORIBA clinical chemistry analyzers.

Intended Use ^a

ABX Pentra Etching CP is a solution used to clean the sodium electrode of the ISE module of the HORIBA clinical chemistry analysers. Clinical laboratories use.

Reagents

ABX Pentra Etching CP is ready-to-use. This reagent contains a solution of ammonium fluoride.

ABX Pentra Etching CP should be used according to this notice. The manufacturer cannot guarantee its performance if used otherwise.

Handling

1. Remove the cap of the cassette.
2. Place **ABX Pentra Etching CP** on the instrument:
 - For **Pentra C200**: Fill a tube with **ABX Pentra Etching CP**. Place the tube in the sample tray.
 - For **Pentra C400**: Place the cassette in the instrument reagent compartment at room temperature position.
 - For **ABX Pentra 400**: Place the cassette in the instrument reagent compartment at room temperature position.

For further details, please refer to the maintenance procedures outlined in the User Manual for Pentra C200 or ABX Pentra 400 or Pentra C400 clinical chemistry analyzers.

Materials required but not provided

- Automated clinical chemistry analyzer equipped with ISE module option.
- Standard laboratory equipment.

Storage and Stability

Stability before opening:

Stable up to the expiry date on the label if stored at 15-25°C.

Stability after opening:

Stable for 1 month at 15-25°C if closed immediately and contamination is avoided.

Waste Management

Please refer to local legal requirements.

General Precautions ^b

- This reagent is for professional *in vitro* diagnostic use only. For laboratory use.
- For prescription use only.
- This reagent is classified as hazardous in compliance with regulation (EC) N°.1272/2008.

^aModification: new leaflet form.

^bModification: modification of classification.

ABX Pentra Etching CP

■ Warning

H315: Causes skin irritation.

H319: Cause serious eye irritation.

P264: Wash hands thoroughly after handling.

P280: Wear protective gloves or clothing and eye or face protection.

P302 + P352: IF ON SKIN: Wash with plenty of soap and water.

P332 + P313: If skin irritation occurs: Get medical advice/attention.

P362 + P364: Take off contaminated clothing and wash it before reuse.

P305 + P351 + P338: IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.

P337 + P313: If eye irritation persists: Get medical advice/attention.

- Do not swallow. Avoid contact with skin and mucous membranes.
- Observe the standard laboratory precautions for use.
- The reagent cassettes are disposable and should be disposed of in accordance with the local legal requirements.
- Please refer to the SDS associated with the reagent.
- Do not use the product if there is visible evidence of biological, chemical or physical deterioration.
- Do not use the product if the recommended storage conditions, including temperature, are not followed.
- User must be trained by a HORIBA representative before attempting to operate the device.
- It is the user's responsibility to verify that this document is applicable to the reagent used.
- For technical assistance, you can call +33 (0)4 67 14 15 16.
- Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the country in which the user and/or the patient is established.