



Yumizen G CTRL I (level 1) for Yumizen G Line



D 4 9 5 0 2 2 1 4 6 4 1 2 0 2 2 7 D



D 5 1 0 3 1 4 0 0 7 9 1 3 1 0 7 8 D



D 6 1 1 7 0 9 6 1 2 9 0 9 3 1 5 5 D



D 7 0 7 2 1 0 9 0 8 0 1 0 8 0 9 7 D



D 8 1 6 2 0 7 3 1 0 9 3 4 5 4 6 7 D



D 9 3 1 7 4 2 9 2 0 3 3 0 4 2 1 6 D



D 0 3 2 4 3 5 6 D

Reference No.:		1300036412	
Lot No:		950221 (rev. 3)	
Expiry date:		2027-02-28	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	10,3-14,0
	Prothrombin %	%	79-131
	Prothrombin INR	INR	0,78-1,17
PT Liq	Prothrombin Time	sec	9,6-12,9
	Prothrombin %	%	93-155
	Prothrombin INR	INR	0,72-1,09
PT R	Prothrombin Time	sec	8,0-10,8
	Prothrombin %	%	97-162
	Prothrombin INR	INR	0,73-1,09
APTT	Partial Thromboplastin Time	sec	34,5-46,7
APTT Liq	Partial Thromboplastin Time	sec	31,7-42,9
FIB Cl	Fibrinogen	g/L	2,03-3,04
TT	Thrombin Time	sec	21,6-32,4
AT	Antithrombin	%	79-107

Important notice

The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.

The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.



Yumizen G CTRL II (level 2) for Yumizen G Line



D 4 9 5 0 2 2 1 5 6 4 1 2 0 2 2 7 D



D 5 2 0 0 3 0 0 0 2 3 0 3 8 1 8 3 D



D 6 2 7 4 1 7 9 2 6 9 0 2 5 0 4 2 D



D 7 1 7 1 2 5 7 1 5 3 2 3 0 0 2 4 D



D 8 0 3 9 1 5 3 2 3 0 6 8 6 9 2 8 D



D 9 7 0 2 9 5 0 0 6 8 1 5 9 0 0 0 D



D 0 9 9 9 3 6 7 D

Reference No.:		1300036412	
Lot No:		950221 (rev. 3)	
Expiry date:		2027-02-28	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	20,0-30,0
	Prothrombin %	%	23-38
	Prothrombin INR	INR	1,83-2,74
PT Liq	Prothrombin Time	sec	17,9-26,9
	Prothrombin %	%	25-42
	Prothrombin INR	INR	1,71-2,57
PT R	Prothrombin Time	sec	15,3-23,0
	Prothrombin %	%	24-39
	Prothrombin INR	INR	1,53-2,30
APTT	Partial Thromboplastin Time	sec	68,6-92,8
APTT Liq	Partial Thromboplastin Time	sec	70,2-95,0
FIB Cl	Fibrinogen	g/L	0,68-1,59
TT	Thrombin Time	sec	N/A
AT	Antithrombin	%	34-57
Important notice			
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.</p>			