



# Yumizen G CTRL I (level 1) for Yumizen G Line



D 4 9 5 0 6 1 8 4 6 4 1 2 0 6 2 7 D



D 5 1 1 3 1 5 3 0 6 5 1 0 8 0 8 7 D



D 6 1 3 0 1 0 6 1 4 3 0 7 1 1 1 9 D



D 7 0 8 2 1 2 3 0 8 9 1 2 1 0 7 5 D



D 8 1 2 5 0 8 0 1 2 0 3 4 0 4 6 0 D



D 9 3 0 4 4 1 2 1 7 4 2 6 0 2 0 8 D



D 0 3 1 2 3 2 2 D

Reference No.:		1300036412	
Lot No:		950618 (rev. 1)	
Expiry date:		2027-06-30	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	11,3-15,3
	Prothrombin %	%	65-108
	Prothrombin INR	INR	0,87-1,30
PT Liq	Prothrombin Time	sec	10,6-14,3
	Prothrombin %	%	71-119
	Prothrombin INR	INR	0,82-1,23
PT R	Prothrombin Time	sec	8,9-12,1
	Prothrombin %	%	75-125
	Prothrombin INR	INR	0,80-1,20
APTT	Partial Thromboplastin Time	sec	34,0-46,0
APTT Liq	Partial Thromboplastin Time	sec	30,4-41,2
FIB Cl	Fibrinogen	g/L	1,74-2,60
TT	Thrombin Time	sec	20,8-31,2
AT	Antithrombin	%	82-111

### Important notice

The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.

The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.



# Yumizen G CTRL II (level 2) for Yumizen G Line



D 4 9 5 0 6 1 8 5 6 4 1 2 0 6 2 7 D



D 5 1 8 5 2 7 8 0 2 6 0 4 3 1 6 7 D



D 6 2 5 1 1 7 6 2 6 3 0 2 6 0 4 3 D



D 7 1 6 6 2 5 0 1 3 9 2 0 8 0 2 8 D



D 8 0 4 6 1 3 4 2 0 1 6 9 7 9 4 2 D



D 9 7 0 2 9 5 0 0 6 7 1 5 6 0 0 0 D



D 0 9 9 9 3 9 5 D

Reference No.:		1300036412	
Lot No:		950618 (rev. 1)	
Expiry date:		2027-06-30	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	18,5-27,8
	Prothrombin %	%	26-43
	Prothrombin INR	INR	1,67-2,51
PT Liq	Prothrombin Time	sec	17,6-26,3
	Prothrombin %	%	26-43
	Prothrombin INR	INR	1,66-2,50
PT R	Prothrombin Time	sec	13,9-20,8
	Prothrombin %	%	28-46
	Prothrombin INR	INR	1,34-2,01
APTT	Partial Thromboplastin Time	sec	69,7-94,2
APTT Liq	Partial Thromboplastin Time	sec	70,2-95,0
FIB Cl	Fibrinogen	g/L	0,67-1,56
TT	Thrombin Time	sec	NA
AT	Antithrombin	%	34-57
<b>Important notice</b>			
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.</p>			