

**HORIBA ABX SAS**  
 Parc Euromédecine  
 Rue du Caducée  
 B.P. 7290  
 34184 Montpellier Cedex 4  
 FRANCE  
 www.horiba.com



**LOT** 25023 (rev. 01)



2 5 0 2 3

**REF** :  
 1300023947

2026/09/30

**SAP** 1300029568



1300029568

### For instruments ABX Pentra 400 – Pentra C400 – Pentra C200

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	ABX Pentra 400 (P400)	Pentra C400 (PC400)	Pentra C200 (PC200)	Unité Unit Einheit Unidad Unità Unidade Enhed Enhet Μονάδα Jednostka Enhet	Valeur titrée Assigned value Deklariertes Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tildelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd	Intervalle de confiance Confidence range Vertrauensintervalle Intervalos de confianza Livelli di fiducia Intervalo de confiança Konfidensområde Konfidensintervall Εύρος εμπιστοσύνης Zakres ufności Konfidensspekter	
				C	U	⊕	Confidence range*
AlbT2-U ABX Pentra Micro ALBUMIN 2 CP (1300032563)	•			mg/L mg/dL	<b>64.4</b> <b>6.44</b>	51.6 - 77.2 5.16 - 7.72	6.4 0.64
		•		mg/L mg/dL	<b>64.4</b> <b>6.44</b>	51.6 - 77.2 5.16 - 7.72	6.4 0.64
			•	mg/L mg/dL	<b>62.9</b> <b>6.29</b>	50.3 - 75.5 5.03 - 7.55	6.3 0.63
AMY (EPS-G7) ABX Pentra Amylase CP (A11A01628)	•			U/L μkat/L	<b>247.1</b> <b>4.127</b>	197.7 - 296.5 3.303 - 4.951	24.7 0.412
		•		U/L μkat/L	<b>247.1</b> <b>4.127</b>	197.7 - 296.5 3.303 - 4.951	24.7 0.412
			•	U/L μkat/L	<b>241.8</b> <b>4.038</b>	193.4 - 290.2 3.230 - 4.846	24.2 0.404
CA (AZ III) ABX Pentra Calcium AS CP (A11A01954)	•			mmol/L mg/L mg/dL	<b>3.02</b> <b>121.1</b> <b>12.11</b>	2.42 - 3.62 97.1 - 145.1 9.71 - 14.51	0.30 12.0 1.20
		•		mmol/L mg/L mg/dL	<b>3.02</b> <b>121.1</b> <b>12.11</b>	2.42 - 3.62 97.1 - 145.1 9.71 - 14.51	0.30 12.0 1.20
			•	mmol/L mg/L mg/dL	<b>2.96</b> <b>118.7</b> <b>11.87</b>	2.36 - 3.56 94.7 - 142.7 9.47 - 14.27	0.30 12.0 1.20
				mmol/L mg/L mg/dL	<b>3.02</b> <b>121.1</b> <b>12.11</b>	2.42 - 3.62 97.1 - 145.1 9.71 - 14.51	0.30 12.0 1.20
		•		mmol/L mg/L mg/dL	<b>3.02</b> <b>121.1</b> <b>12.11</b>	2.42 - 3.62 97.1 - 145.1 9.71 - 14.51	0.30 12.0 1.20
			•	mmol/L mg/L mg/dL	<b>2.96</b> <b>118.7</b> <b>11.87</b>	2.36 - 3.56 94.7 - 142.7 9.47 - 14.27	0.30 12.0 1.20

\* Confidence range: +/- 2 SD

**For instruments ABX Pentra 400 – Pentra C400 – Pentra C200**

C	P400	PC400	PC200	U	⊕	Confidence range*	1 SD		
CREA (CREA3 / CREA4) (Jaffe/Alkaline Picrate) ABX Pentra Creatinine 120 CP (A11A01933)	•			μmol/L mmol/L mg/dL	<b>13123</b> <b>13.12</b> <b>148.3</b>	10499 - 15747 10.50 - 15.74 118.7 - 177.9	1312 1.31 14.8		
		•		μmol/L mmol/L mg/dL	<b>13123</b> <b>13.12</b> <b>148.3</b>	10499 - 15747 10.50 - 15.74 118.7 - 177.9	1312 1.31 14.8		
			•	μmol/L mmol/L mg/dL	<b>12282</b> <b>12.28</b> <b>138.8</b>	9826 - 14738 9.82 - 14.74 111.0 - 166.6	1228 1.23 13.9		
	CREA ENZ. ABX Pentra Enzymatic Creatinine CP (A11A01907)	•			μmol/L mmol/L mg/dL	<b>14706</b> <b>14.71</b> <b>166.2</b>	11764 - 17648 11.77 - 17.65 133.0 - 199.4	1471 1.47 16.6	
			•		μmol/L mmol/L mg/dL	<b>14706</b> <b>14.71</b> <b>166.2</b>	11764 - 17648 11.77 - 17.65 133.0 - 199.4	1471 1.47 16.6	
				•	μmol/L mmol/L mg/dL	<b>14066</b> <b>14.07</b> <b>158.9</b>	11252 - 16880 11.25 - 16.89 127.1 - 190.7	1407 1.41 15.9	
		GluK (HK) ABX Pentra Glucose HK CP (A11A01667)	•			mmol/L g/L mg/dL	<b>15.72</b> <b>2.830</b> <b>283.0</b>	12.58 - 18.86 2.264 - 3.396 226.4 - 339.6	1.57 0.283 28.3
				•		mmol/L g/L mg/dL	<b>15.72</b> <b>2.830</b> <b>283.0</b>	12.58 - 18.86 2.264 - 3.396 226.4 - 339.6	1.57 0.283 28.3
					•	mmol/L g/L mg/dL	<b>15.65</b> <b>2.817</b> <b>281.7</b>	12.51 - 18.79 2.251 - 3.383 225.1 - 338.3	1.57 0.283 28.3
GluP (PAP) ABX Pentra Glucose PAP CP (A11A01668)	•			mmol/L g/L mg/dL	<b>17.66</b> <b>3.179</b> <b>317.9</b>	14.12 - 21.20 2.541 - 3.817 254.1 - 381.7	1.77 0.319 31.9		
		•		mmol/L g/L mg/dL	<b>17.66</b> <b>3.179</b> <b>317.9</b>	14.12 - 21.20 2.541 - 3.817 254.1 - 381.7	1.77 0.319 31.9		
			•	mmol/L g/L mg/dL	<b>16.35</b> <b>2.943</b> <b>294.3</b>	13.07 - 19.63 2.353 - 3.533 235.3 - 353.3	1.64 0.295 29.5		
	PHOS (Ammonium Molybdate UV) ABX Pentra Phosphorus CP (A11A01665)	•			mmol/L mg/L mg/dL	<b>14.29</b> <b>443</b> <b>44.3</b>	11.43 - 17.15 355 - 531 35.5 - 53.1	1.43 44 4.4	
			•		mmol/L mg/L mg/dL	<b>14.29</b> <b>443</b> <b>44.3</b>	11.43 - 17.15 355 - 531 35.5 - 53.1	1.43 44 4.4	
				•	mmol/L mg/L mg/dL	<b>13.09</b> <b>406</b> <b>40.6</b>	10.47 - 15.71 324 - 488 32.4 - 48.8	1.31 41 4.1	

\* Confidence range: +/- 2 SD

**For instruments ABX Pentra 400 – Pentra C400 – Pentra C200**

C	P400	PC400	PC200	U	⊕	Confidence range*	1 SD
TPU (Pyrogallol red) ABX Pentra Urinary Proteins CP (A11A01642)	•			g/L	<b>0.576</b>	0.460 - 0.692	0.058
				mg/L	<b>576</b>	460 - 692	58
				mg/dL	<b>57.6</b>	46.0 - 69.2	5.8
		•		g/L	<b>0.576</b>	0.460 - 0.692	0.058
				mg/L	<b>576</b>	460 - 692	58
				mg/dL	<b>57.6</b>	46.0 - 69.2	5.8
UREA (GLDH) ABX Pentra Urea CP (A11A01641)			•	g/L	<b>0.574</b>	0.460 - 0.688	0.057
				mg/L	<b>574</b>	460 - 688	57
				mg/dL	<b>57.4</b>	46.0 - 68.8	5.7
	•			mmol/L	<b>283</b>	227 - 339	28
				mg/L	<b>16980</b>	13620 - 20340	1680
				mg/dL	<b>1698</b>	1362 - 2034	168
BUN (GLDH) ABX Pentra Urea CP (A11A01641)		•		mmol/L	<b>283</b>	227 - 339	28
				mg/L	<b>16980</b>	13620 - 20340	1680
				mg/dL	<b>1698</b>	1362 - 2034	168
			•	mmol/L	<b>257</b>	205 - 309	26
				mg/L	<b>15420</b>	12300 - 18540	1560
				mg/dL	<b>1542</b>	1230 - 1854	156
UA (Uricase/Peroxidase) ABX Pentra Uric Acid CP (A11A01670)	•			mmol/L	<b>283</b>	227 - 339	28
				mg/dL	<b>795</b>	637 - 953	79
		•		mmol/L	<b>283</b>	227 - 339	28
				mg/dL	<b>795</b>	637 - 953	79
			•	mmol/L	<b>257</b>	205 - 309	26
				mg/dL	<b>722</b>	576 - 868	73
UA (Uricase/Peroxidase) ABX Pentra Uric Acid CP (A11A01670)	•			μmol/L	<b>829</b>	663 - 995	83
				mmol/L	<b>0.83</b>	0.67 - 0.99	0.08
				mg/dL	<b>13.93</b>	11.15 - 16.71	1.39
		•		μmol/L	<b>829</b>	663 - 995	83
				mmol/L	<b>0.83</b>	0.67 - 0.99	0.08
				mg/dL	<b>13.93</b>	11.15 - 16.71	1.39
UA (Uricase/Peroxidase) ABX Pentra Uric Acid CP (A11A01670)			•	μmol/L	<b>825</b>	659 - 991	83
				mmol/L	<b>0.83</b>	0.67 - 0.99	0.08
				mg/dL	<b>13.86</b>	11.08 - 16.64	1.39

\* Confidence range: +/- 2 SD