

HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée
 B.P. 7290
 34184 Montpellier Cedex 4
 FRANCE
 www.horiba.com



LOT 25041 (rev. 01)



2 5 0 4 1

REF :
 1300023947

2026/09/30

SAP 1300029568



1300029568

For instruments ABX Pentra 400 – Pentra C400 – Pentra C200

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	ABX Pentra 400 (P400)	Pentra C400 (PC400)	Pentra C200 (PC200)	Unité Unit Einheit Unidad Unità Unidade Enhed Enhet Μονάδα Jednostka Enhet	Valeur titrée Assigned value Deklariertes Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tildelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd	Intervalle de confiance Confidence range Vertrauensintervalle Intervalos de confianza Livelli di fiducia Intervalo de confiança Konfidensområde Konfidensintervall Εύρος εμπιστοσύνης Zakres ufnosci Konfidensspekter	
C				U	\oplus	Confidence range*	1 SD
AlbT2-U ABX Pentra Micro ALBUMIN 2 CP (1300032563)	•			mg/L mg/dL	64.4 6.44	51.6 - 77.2 5.16 - 7.72	6.4 0.64
		•		mg/L mg/dL	64.4 6.44	51.6 - 77.2 5.16 - 7.72	6.4 0.64
			•	mg/L mg/dL	62.9 6.29	50.3 - 75.5 5.03 - 7.55	6.3 0.63
AMY (EPS-G7) ABX Pentra Amylase CP (A11A01628)	•			U/L μkat/L	247.1 4.127	197.7 - 296.5 3.303 - 4.951	24.7 0.412
		•		U/L μkat/L	247.1 4.127	197.7 - 296.5 3.303 - 4.951	24.7 0.412
			•	U/L μkat/L	241.8 4.038	193.4 - 290.2 3.230 - 4.846	24.2 0.404
CA (AZ III) ABX Pentra Calcium AS CP (A11A01954)	•			mmol/L mg/L mg/dL	3.02 121.1 12.11	2.42 - 3.62 97.1 - 145.1 9.71 - 14.51	0.30 12.0 1.20
		•		mmol/L mg/L mg/dL	3.02 121.1 12.11	2.42 - 3.62 97.1 - 145.1 9.71 - 14.51	0.30 12.0 1.20
			•	mmol/L mg/L mg/dL	2.96 118.7 11.87	2.36 - 3.56 94.7 - 142.7 9.47 - 14.27	0.30 12.0 1.20
				mmol/L	2.96	2.36 - 3.56	0.30
				mg/L	118.7	94.7 - 142.7	12.0
				mg/dL	11.87	9.47 - 14.27	1.20

* Confidence range: +/- 2 SD

For instruments ABX Pentra 400 – Pentra C400 – Pentra C200

C	P400	PC400	PC200	U	⊕	Confidence range*	1 SD		
CREA (CREA3 / CREA4) (Jaffe/Alkaline Picrate) ABX Pentra Creatinine 120 CP (A11A01933)	•			μmol/L mmol/L mg/dL	13123 13.12 148.3	10499 - 15747 10.50 - 15.74 118.7 - 177.9	1312 1.31 14.8		
		•		μmol/L mmol/L mg/dL	13123 13.12 148.3	10499 - 15747 10.50 - 15.74 118.7 - 177.9	1312 1.31 14.8		
			•	μmol/L mmol/L mg/dL	12282 12.28 138.8	9826 - 14738 9.82 - 14.74 111.0 - 166.6	1228 1.23 13.9		
	CREA ENZ. ABX Pentra Enzymatic Creatinine CP (A11A01907)	•			μmol/L mmol/L mg/dL	14706 14.71 166.2	11764 - 17648 11.77 - 17.65 133.0 - 199.4	1471 1.47 16.6	
			•		μmol/L mmol/L mg/dL	14706 14.71 166.2	11764 - 17648 11.77 - 17.65 133.0 - 199.4	1471 1.47 16.6	
				•	μmol/L mmol/L mg/dL	14066 14.07 158.9	11252 - 16880 11.25 - 16.89 127.1 - 190.7	1407 1.41 15.9	
		GluK (HK) ABX Pentra Glucose HK CP (A11A01667)	•			mmol/L g/L mg/dL	15.72 2.830 283.0	12.58 - 18.86 2.264 - 3.396 226.4 - 339.6	1.57 0.283 28.3
				•		mmol/L g/L mg/dL	15.72 2.830 283.0	12.58 - 18.86 2.264 - 3.396 226.4 - 339.6	1.57 0.283 28.3
					•	mmol/L g/L mg/dL	15.65 2.817 281.7	12.51 - 18.79 2.251 - 3.383 225.1 - 338.3	1.57 0.283 28.3
GluP (PAP) ABX Pentra Glucose PAP CP (A11A01668)	•			mmol/L g/L mg/dL	17.66 3.179 317.9	14.12 - 21.20 2.541 - 3.817 254.1 - 381.7	1.77 0.319 31.9		
		•		mmol/L g/L mg/dL	17.66 3.179 317.9	14.12 - 21.20 2.541 - 3.817 254.1 - 381.7	1.77 0.319 31.9		
			•	mmol/L g/L mg/dL	16.35 2.943 294.3	13.07 - 19.63 2.353 - 3.533 235.3 - 353.3	1.64 0.295 29.5		
	PHOS (Ammonium Molybdate UV) ABX Pentra Phosphorus CP (A11A01665)	•			mmol/L mg/L mg/dL	14.29 443 44.3	11.43 - 17.15 355 - 531 35.5 - 53.1	1.43 44 4.4	
			•		mmol/L mg/L mg/dL	14.29 443 44.3	11.43 - 17.15 355 - 531 35.5 - 53.1	1.43 44 4.4	
				•	mmol/L mg/L mg/dL	13.09 406 40.6	10.47 - 15.71 324 - 488 32.4 - 48.8	1.31 41 4.1	

* Confidence range: +/- 2 SD

For instruments ABX Pentra 400 – Pentra C400 – Pentra C200

C	P400	PC400	PC200	U	⊕	Confidence range*	1 SD
TPU (Pyrogallol red) ABX Pentra Urinary Proteins CP (A11A01642)	•			g/L	0.576	0.460 - 0.692	0.058
				mg/L	576	460 - 692	58
				mg/dL	57.6	46.0 - 69.2	5.8
		•		g/L	0.576	0.460 - 0.692	0.058
				mg/L	576	460 - 692	58
				mg/dL	57.6	46.0 - 69.2	5.8
UREA (GLDH) ABX Pentra Urea CP (A11A01641)			•	g/L	0.574	0.460 - 0.688	0.057
				mg/L	574	460 - 688	57
				mg/dL	57.4	46.0 - 68.8	5.7
	•			mmol/L	283	227 - 339	28
				mg/L	16980	13620 - 20340	1680
				mg/dL	1698	1362 - 2034	168
BUN (GLDH) ABX Pentra Urea CP (A11A01641)		•		mmol/L	283	227 - 339	28
				mg/L	16980	13620 - 20340	1680
				mg/dL	1698	1362 - 2034	168
			•	mmol/L	257	205 - 309	26
				mg/L	15420	12300 - 18540	1560
				mg/dL	1542	1230 - 1854	156
UA (Uricase/Peroxidase) ABX Pentra Uric Acid CP (A11A01670)	•			mmol/L	283	227 - 339	28
				mg/dL	795	637 - 953	79
		•		mmol/L	283	227 - 339	28
				mg/dL	795	637 - 953	79
			•	mmol/L	257	205 - 309	26
				mg/dL	722	576 - 868	73
UA (Uricase/Peroxidase) ABX Pentra Uric Acid CP (A11A01670)	•			μmol/L	829	663 - 995	83
				mmol/L	0.83	0.67 - 0.99	0.08
				mg/dL	13.93	11.15 - 16.71	1.39
		•		μmol/L	829	663 - 995	83
				mmol/L	0.83	0.67 - 0.99	0.08
				mg/dL	13.93	11.15 - 16.71	1.39
UA (Uricase/Peroxidase) ABX Pentra Uric Acid CP (A11A01670)			•	μmol/L	825	659 - 991	83
				mmol/L	0.83	0.67 - 0.99	0.08
				mg/dL	13.86	11.08 - 16.64	1.39

* Confidence range: +/- 2 SD