



Yumizen G CAL

for Yumizen G Line

Reference No.:		1300036416
Lot No.:		950504 (rev. 1)
Expiry date:		2027-05-31
Parameter	Unit	Value
PT	%	89
PT Liq	%	89
PT R	%	83
Fibrinogen	g/L	2.95
	mg/dL	295
Antithrombin	%	103
	IU/mL	1,03
Factor II Deficient Plasma	%	N/A
	IU/mL	N/A
Factor V Deficient Plasma	%	N/A
	IU/mL	N/A
Factor VII Deficient Plasma	%	N/A
	IU/mL	N/A
Factor VIII Deficient Plasma	%	N/A
	IU/mL	N/A
Factor IX Deficient Plasma	%	N/A
	IU/mL	N/A
Factor X Deficient Plasma	%	N/A
	IU/mL	N/A
Factor XI Deficient Plasma	%	N/A
	IU/mL	N/A
Factor XII Deficient Plasma	%	N/A
	IU/mL	N/A
Important notice		
Dilution serie must be prepared. Please refer to the notice.		