

LOT

CX 512

CAL



(Exp.) 2025-12-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		ES60				60 8P 60 18P					
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.0				9.2					± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.34				4.29					± 0.06
HB HGB	g/dl	12.8				12.7					± 0.2
	g/l	128				127					± 2
	mmol/l	7.95				7.89					± 0.12
HT HCT	%	36.0				36.3					± 1.0
	l/l	0.360				0.363					± 0.010
PLA PLT	10 ³ /mm ³ ;10 ⁹ /l	225				238					± 10
VMP MPV	μm ³ ; fl	8.1				7.9					± 0.5

ABX Minocal

LOT

CX 512

CAL



(Exp.) 2025-12-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	LyseVet								TOLERANCES TOLERANCE
		MICROS	MICROS							
			Scil Vet ABC Plus Scil Vet ABC ESV 60							
GB WBC	10 ³ /mm ³ ;10 ⁹ /l		8.8							± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l		4.32							± 0.06
HB HGB	g/dl		12.5							± 0.2
	g/l		125							± 2
	mmol/l		7.76							± 0.12
HT HCT	%		35.9							± 1.0
	l/l		0.359							± 0.010
PLA PLT	10 ³ /mm ³ ;10 ⁹ /l		233							± 10
VMP MPV *	µm ³ ; fl		7.8							± 0.5

Ref: TEMP-0387 Rev.53

* N/A for Micros Scil Vet ABC