



# Yumizen G CTRL I (level 1) for Yumizen G Line



D4950526464120527D



D5111150069114085D



D6127105143072120D



D7082123093126068D



D8113083125353478D



D9286387215323209D



D0313349D

Reference No.:		1300036412	
Lot No:		950526 (rev. 1)	
Expiry date:		2027-05-31	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	11,1-15,0
	Prothrombin %	%	69-114
	Prothrombin INR	INR	0,85-1,27
PT Liq	Prothrombin Time	sec	10,5-14,3
	Prothrombin %	%	72-120
	Prothrombin INR	INR	0,82-1,23
PT R	Prothrombin Time	sec	9,3-12,6
	Prothrombin %	%	68-113
	Prothrombin INR	INR	0,83-1,25
APTT	Partial Thromboplastin Time	sec	35,3-47,8
APTT Liq	Partial Thromboplastin Time	sec	28,6-38,7
FIB Cl	Fibrinogen	g/L	2,15-3,23
TT	Thrombin Time	sec	20,9-31,3
AT	Antithrombin	%	84-114

### Important notice

The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.

The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.



# Yumizen G CTRL II (level 2) for Yumizen G Line



Reference No.:		1300036412	
Lot No:		950526 (rev. 1)	
Expiry date:		2027-05-31	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	17,3-26,0
	Prothrombin %	%	28-47
	Prothrombin INR	INR	1,54-2,31
PT Liq	Prothrombin Time	sec	16,5-24,7
	Prothrombin %	%	28-47
	Prothrombin INR	INR	1,54-2,31
PT R	Prothrombin Time	sec	14,2-21,2
	Prothrombin %	%	27-45
	Prothrombin INR	INR	1,37-2,05
APTT	Partial Thromboplastin Time	sec	60,2-81,5
APTT Liq	Partial Thromboplastin Time	sec	59,1-80,0
FIB Cl	Fibrinogen	g/L	0,72-1,68
TT	Thrombin Time	sec	N/A
AT	Antithrombin	%	36-59
<b>Important notice</b>			
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.</p>			